

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: Core Environmental Consultants and Services, Inc.

SDG No.: Q2681

Contract: CORE02

Lab Code: ACE

ICS Source: EPA

Instrument ID: P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Arsenic	9.84			-20	20	07/28/2025	12:11	LB136640
	Barium	5.33	6.0	89	-94	106	07/28/2025	12:11	LB136640
	Cadmium	5.35	1.0	535	-5	7	07/28/2025	12:11	LB136640
	Chromium	55.1	52.0	106	42	62	07/28/2025	12:11	LB136640
	Lead	-1.40			-12	12	07/28/2025	12:11	LB136640
	Selenium	-5.30			-20	20	07/28/2025	12:11	LB136640
	Silver	2.18			-10	10	07/28/2025	12:11	LB136640
ICSAB01	Arsenic	102	104	98	88.4	120	07/28/2025	12:16	LB136640
	Barium	498	537	93	437	637	07/28/2025	12:16	LB136640
	Cadmium	913	972	94	826	1120	07/28/2025	12:16	LB136640
	Chromium	531	542	98	460	624	07/28/2025	12:16	LB136640
	Lead	39.0	49.0	80	37	61	07/28/2025	12:16	LB136640
	Selenium	49.0	46.0	106	26	66	07/28/2025	12:16	LB136640
	Silver	213	201	106	170	232	07/28/2025	12:16	LB136640