

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client:	<u>Core Environmental Consultants and Service</u>	SDG No.:	<u>Q2681</u>
Contract:	<u>CORE02</u>	Lab Code:	<u>ACE</u>

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB169012BS							
Arsenic	ug/L	4000	3530		88	80 - 120	P
Barium	ug/L	1000	1060		106	80 - 120	P
Cadmium	ug/L	1000	840		84	80 - 120	P
Chromium	ug/L	2000	1890		94	80 - 120	P
Lead	ug/L	5000	4310		86	80 - 120	P
Selenium	ug/L	10000	9000		90	80 - 120	P
Silver	ug/L	380	359		94	80 - 120	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Core Environmental Consultants and Service

SDG No.: Q2681

Contract: CORE02

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB169026BS Mercury	ug/L	4.0	4.72		118	80 - 120	CV