

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** CORE02
Lab Code: ACE **SDG NO.:** Q2681
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/23/2025 07/23/2025

Client Sample No.: CCAL01 **Date Analyzed:** 07/24/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO112479.D **Time Analyzed:** 20:46

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.756	4.655	4.855	542.590	500.000	8.5
Aroclor-1016-2	4.774	4.674	4.874	533.810	500.000	6.8
Aroclor-1016-3	4.831	4.731	4.931	521.550	500.000	4.3
Aroclor-1016-4	4.951	4.851	5.051	518.050	500.000	3.6
Aroclor-1016-5	5.208	5.108	5.308	568.320	500.000	13.7
Aroclor-1260-1	6.245	6.145	6.345	497.050	500.000	-0.6
Aroclor-1260-2	6.434	6.333	6.533	485.400	500.000	-2.9
Aroclor-1260-3	6.800	6.700	6.900	508.790	500.000	1.8
Aroclor-1260-4	7.059	6.959	7.159	520.010	500.000	4.0
Aroclor-1260-5	7.302	7.203	7.403	484.470	500.000	-3.1
Decachlorobiphenyl	8.694	8.593	8.793	43.150	50.000	-13.7
Tetrachloro-m-xylene	3.670	3.569	3.769	55.090	50.000	10.2

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** CORE02
Lab Code: ACE **SDG NO.:** Q2681
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/23/2025 07/23/2025

Client Sample No.: CCAL01 **Date Analyzed:** 07/24/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO112479.D **Time Analyzed:** 20:46

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.738	4.638	4.838	520.250	500.000	4.1
Aroclor-1016-2	4.756	4.656	4.856	520.700	500.000	4.1
Aroclor-1016-3	4.931	4.831	5.031	523.770	500.000	4.8
Aroclor-1016-4	4.973	4.873	5.073	518.760	500.000	3.8
Aroclor-1016-5	5.185	5.085	5.285	543.110	500.000	8.6
Aroclor-1260-1	6.214	6.113	6.313	519.010	500.000	3.8
Aroclor-1260-2	6.402	6.302	6.502	509.010	500.000	1.8
Aroclor-1260-3	6.553	6.453	6.653	520.260	500.000	4.1
Aroclor-1260-4	7.023	6.923	7.123	512.310	500.000	2.5
Aroclor-1260-5	7.265	7.164	7.364	522.080	500.000	4.4
Decachlorobiphenyl	8.641	8.540	8.740	47.110	50.000	-5.8
Tetrachloro-m-xylene	3.663	3.562	3.762	54.170	50.000	8.3