

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** CORE02
Lab Code: ACE **SDG NO.:** Q2681
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/23/2025 07/23/2025

Client Sample No.: CCAL02 **Date Analyzed:** 07/25/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO112494.D **Time Analyzed:** 05:52

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.754	4.655	4.855	565.380	500.000	13.1
Aroclor-1016-2	4.773	4.674	4.874	560.270	500.000	12.1
Aroclor-1016-3	4.830	4.731	4.931	549.370	500.000	9.9
Aroclor-1016-4	4.950	4.851	5.051	551.370	500.000	10.3
Aroclor-1016-5	5.206	5.108	5.308	558.510	500.000	11.7
Aroclor-1260-1	6.244	6.145	6.345	366.480	500.000	-26.7
Aroclor-1260-2	6.433	6.333	6.533	362.490	500.000	-27.5
Aroclor-1260-3	6.800	6.700	6.900	594.190	500.000	18.8
Aroclor-1260-4	7.058	6.959	7.159	468.900	500.000	-6.2
Aroclor-1260-5	7.302	7.203	7.403	356.920	500.000	-28.6
Decachlorobiphenyl	8.694	8.593	8.793	33.470	50.000	-33.1
Tetrachloro-m-xylene	3.669	3.569	3.769	56.980	50.000	14.0

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** CORE02
Lab Code: ACE **SDG NO.:** Q2681
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/23/2025 07/23/2025

Client Sample No.: CCAL02 **Date Analyzed:** 07/25/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO112494.D **Time Analyzed:** 05:52

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.738	4.638	4.838	531.680	500.000	6.3
Aroclor-1016-2	4.756	4.656	4.856	528.680	500.000	5.7
Aroclor-1016-3	4.931	4.831	5.031	537.510	500.000	7.5
Aroclor-1016-4	4.973	4.873	5.073	521.330	500.000	4.3
Aroclor-1016-5	5.185	5.085	5.285	583.170	500.000	16.6
Aroclor-1260-1	6.213	6.113	6.313	481.020	500.000	-3.8
Aroclor-1260-2	6.402	6.302	6.502	442.420	500.000	-11.5
Aroclor-1260-3	6.553	6.453	6.653	449.420	500.000	-10.1
Aroclor-1260-4	7.023	6.923	7.123	430.010	500.000	-14.0
Aroclor-1260-5	7.265	7.164	7.364	437.430	500.000	-12.5
Decachlorobiphenyl	8.641	8.540	8.740	42.590	50.000	-14.8
Tetrachloro-m-xylene	3.663	3.562	3.762	55.140	50.000	10.3