

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** CORE02
Lab Code: ACE **SDG NO.:** Q2681
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/23/2025 07/23/2025

Client Sample No.: CCAL03 **Date Analyzed:** 07/25/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO112499.D **Time Analyzed:** 09:45

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.759	4.655	4.855	549.190	500.000	9.8
Aroclor-1016-2	4.778	4.674	4.874	541.360	500.000	8.3
Aroclor-1016-3	4.834	4.731	4.931	532.250	500.000	6.5
Aroclor-1016-4	4.954	4.851	5.051	532.510	500.000	6.5
Aroclor-1016-5	5.210	5.108	5.308	543.600	500.000	8.7
Aroclor-1260-1	6.248	6.145	6.345	523.620	500.000	4.7
Aroclor-1260-2	6.438	6.333	6.533	512.390	500.000	2.5
Aroclor-1260-3	6.804	6.700	6.900	470.070	500.000	-6.0
Aroclor-1260-4	7.063	6.959	7.159	389.290	500.000	-22.1
Aroclor-1260-5	7.306	7.203	7.403	365.030	500.000	-27.0
Decachlorobiphenyl	8.698	8.593	8.793	43.100	50.000	-13.8
Tetrachloro-m-xylene	3.673	3.569	3.769	55.410	50.000	10.8

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** CORE02
Lab Code: ACE **SDG NO.:** Q2681
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/23/2025 07/23/2025

Client Sample No.: CCAL03 **Date Analyzed:** 07/25/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO112499.D **Time Analyzed:** 09:45

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.738	4.638	4.838	522.120	500.000	4.4
Aroclor-1016-2	4.756	4.656	4.856	511.460	500.000	2.3
Aroclor-1016-3	4.931	4.831	5.031	525.990	500.000	5.2
Aroclor-1016-4	4.974	4.873	5.073	513.780	500.000	2.8
Aroclor-1016-5	5.186	5.085	5.285	560.490	500.000	12.1
Aroclor-1260-1	6.215	6.113	6.313	532.970	500.000	6.6
Aroclor-1260-2	6.403	6.302	6.502	523.400	500.000	4.7
Aroclor-1260-3	6.555	6.453	6.653	536.090	500.000	7.2
Aroclor-1260-4	7.024	6.923	7.123	512.400	500.000	2.5
Aroclor-1260-5	7.267	7.164	7.364	510.340	500.000	2.1
Decachlorobiphenyl	8.642	8.540	8.740	54.330	50.000	8.7
Tetrachloro-m-xylene	3.663	3.562	3.762	53.730	50.000	7.5