

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** CORE02
Lab Code: ACE **SDG NO.:** Q2681
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/23/2025 07/23/2025

Client Sample No.: CCAL04 **Date Analyzed:** 07/25/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO112514.D **Time Analyzed:** 15:32

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.756	4.655	4.855	541.980	500.000	8.4
Aroclor-1016-2	4.775	4.674	4.874	535.870	500.000	7.2
Aroclor-1016-3	4.832	4.731	4.931	526.390	500.000	5.3
Aroclor-1016-4	4.952	4.851	5.051	528.380	500.000	5.7
Aroclor-1016-5	5.208	5.108	5.308	520.640	500.000	4.1
Aroclor-1260-1	6.246	6.145	6.345	507.980	500.000	1.6
Aroclor-1260-2	6.436	6.333	6.533	505.850	500.000	1.2
Aroclor-1260-3	6.801	6.700	6.900	526.910	500.000	5.4
Aroclor-1260-4	7.060	6.959	7.159	505.450	500.000	1.1
Aroclor-1260-5	7.304	7.203	7.403	504.390	500.000	0.9
Decachlorobiphenyl	8.695	8.593	8.793	54.450	50.000	8.9
Tetrachloro-m-xylene	3.670	3.569	3.769	54.610	50.000	9.2

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** CORE02
Lab Code: ACE **SDG NO.:** Q2681
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/23/2025 07/23/2025

Client Sample No.: CCAL04 **Date Analyzed:** 07/25/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO112514.D **Time Analyzed:** 15:32

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.738	4.638	4.838	518.090	500.000	3.6
Aroclor-1016-2	4.756	4.656	4.856	521.860	500.000	4.4
Aroclor-1016-3	4.931	4.831	5.031	528.650	500.000	5.7
Aroclor-1016-4	4.974	4.873	5.073	516.190	500.000	3.2
Aroclor-1016-5	5.186	5.085	5.285	560.510	500.000	12.1
Aroclor-1260-1	6.214	6.113	6.313	528.830	500.000	5.8
Aroclor-1260-2	6.402	6.302	6.502	519.860	500.000	4.0
Aroclor-1260-3	6.554	6.453	6.653	531.580	500.000	6.3
Aroclor-1260-4	7.024	6.923	7.123	522.740	500.000	4.5
Aroclor-1260-5	7.266	7.164	7.364	534.360	500.000	6.9
Decachlorobiphenyl	8.642	8.540	8.740	59.450	50.000	18.9
Tetrachloro-m-xylene	3.663	3.562	3.762	54.040	50.000	8.1