

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO072325

Review By	yogesh	Review On	7/23/2025 3:39:31 PM
Supervise By	mohammad	Supervise On	7/25/2025 3:53:10 AM
SubDirectory	PO072325	HP Acquire Method	HP Processing Method PO072325

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP 24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369 PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PO112412.D	23 Jul 2025 10:56		YP/AJ	Ok
2	I.BLK	I.BLK	PO112413.D	23 Jul 2025 11:14		YP/AJ	Ok
3	AR1660ICC1000	AR1660ICC1000	PO112414.D	23 Jul 2025 11:32		YP/AJ	Ok
4	AR1660ICC750	AR1660ICC750	PO112415.D	23 Jul 2025 11:50		YP/AJ	Ok
5	AR1660ICC500	AR1660ICC500	PO112416.D	23 Jul 2025 12:08		YP/AJ	Ok
6	AR1660ICC250	AR1660ICC250	PO112417.D	23 Jul 2025 12:27		YP/AJ	Ok,M
7	AR1660ICC050	AR1660ICC050	PO112418.D	23 Jul 2025 12:45		YP/AJ	Ok,M
8	AR1221ICC500	AR1221ICC500	PO112419.D	23 Jul 2025 13:03		YP/AJ	Ok
9	AR1232ICC500	AR1232ICC500	PO112420.D	23 Jul 2025 13:22		YP/AJ	Ok
10	AR1242ICC1000	AR1242ICC1000	PO112421.D	23 Jul 2025 13:40		YP/AJ	Ok
11	AR1242ICC750	AR1242ICC750	PO112422.D	23 Jul 2025 13:59		YP/AJ	Ok
12	AR1242ICC500	AR1242ICC500	PO112423.D	23 Jul 2025 14:17		YP/AJ	Ok
13	AR1242ICC250	AR1242ICC250	PO112424.D	23 Jul 2025 14:36		YP/AJ	Ok,M
14	AR1242ICC050	AR1242ICC050	PO112425.D	23 Jul 2025 14:54		YP/AJ	Ok,M
15	AR1248ICC1000	AR1248ICC1000	PO112426.D	23 Jul 2025 15:13		YP/AJ	Ok
16	AR1248ICC750	AR1248ICC750	PO112427.D	23 Jul 2025 15:31		YP/AJ	Ok
17	AR1248ICC500	AR1248ICC500	PO112428.D	23 Jul 2025 15:48		YP/AJ	Ok
18	AR1248ICC250	AR1248ICC250	PO112429.D	23 Jul 2025 16:07		YP/AJ	Ok

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO072325

Review By	yogesh	Review On	7/23/2025 3:39:31 PM
Supervise By	mohammad	Supervise On	7/25/2025 3:53:10 AM
SubDirectory	PO072325	HP Acquire Method	HP Processing Method PO072325
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP 24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369 PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387		

19	AR1248ICC050	AR1248ICC050	PO112430.D	23 Jul 2025 16:25		YP/AJ	Ok,M
20	AR1254ICC1000	AR1254ICC1000	PO112431.D	23 Jul 2025 16:44		YP/AJ	Ok
21	AR1254ICC750	AR1254ICC750	PO112432.D	23 Jul 2025 17:02		YP/AJ	Ok
22	AR1254ICC500	AR1254ICC500	PO112433.D	23 Jul 2025 17:21		YP/AJ	Ok
23	AR1254ICC250	AR1254ICC250	PO112434.D	23 Jul 2025 17:39		YP/AJ	Ok
24	AR1254ICC050	AR1254ICC050	PO112435.D	23 Jul 2025 17:57		YP/AJ	Ok,M
25	AR1262ICC500	AR1262ICC500	PO112436.D	23 Jul 2025 18:16		YP/AJ	Ok
26	AR1268ICC1000	AR1268ICC1000	PO112437.D	23 Jul 2025 18:34		YP/AJ	Ok
27	AR1268ICC750	AR1268ICC750	PO112438.D	23 Jul 2025 18:53		YP/AJ	Ok
28	AR1268ICC500	AR1268ICC500	PO112439.D	23 Jul 2025 19:11		YP/AJ	Ok
29	AR1268ICC250	AR1268ICC250	PO112440.D	23 Jul 2025 19:28		YP/AJ	Ok
30	AR1268ICC050	AR1268ICC050	PO112441.D	23 Jul 2025 19:47		YP/AJ	Ok
31	PO072325ICV500	ICVPO072325	PO112442.D	23 Jul 2025 20:05		YP/AJ	Ok
32	AR1242ICV500	ICVPO072325AR1242	PO112443.D	23 Jul 2025 20:42		YP/AJ	Ok
33	AR1248ICV500	ICVPO072325AR1248	PO112444.D	23 Jul 2025 21:18		YP/AJ	Ok
34	AR1254ICV500	ICVPO072325AR1254	PO112445.D	23 Jul 2025 21:54		YP/AJ	Ok
35	AR1268ICV500	ICVPO072325AR1268	PO112446.D	23 Jul 2025 22:31		YP/AJ	Ok
36	DDT ANALOG	DDT ANALOG	PO112447.D	23 Jul 2025 23:07		YP/AJ	Ok

M : Manual Integration

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO072425

Review By	yogesh	Review On	7/24/2025 11:32:48 AM
Supervise By		Supervise On	
SubDirectory	PO072425	HP Acquire Method	HP Processing Method PO070825

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP 24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369
CCC	PP24332,PP24347,PP24352,PP24357
Internal Standard/PEM	
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PO112448.D	24 Jul 2025 08:44		YP/AJ	Ok
2	AR1660CCC500	AR1660CCC500	PO112449.D	24 Jul 2025 09:03		YP/AJ	Ok
3	AR1242CCC500	AR1242CCC500	PO112450.D	24 Jul 2025 09:21		YP/AJ	Ok
4	AR1248CCC500	AR1248CCC500	PO112451.D	24 Jul 2025 09:39		YP/AJ	Ok
5	AR1254CCC500	AR1254CCC500	PO112452.D	24 Jul 2025 09:58		YP/AJ	Ok
6	I.BLK	I.BLK	PO112453.D	24 Jul 2025 10:16		YP/AJ	Ok
7	DDT ANALOG	DDT ANALOG	PO112454.D	24 Jul 2025 10:35		YP/AJ	Ok
8	Q2674-01	TR-04-072225	PO112455.D	24 Jul 2025 10:53		YP/AJ	Ok,NS
9	Q2674-01MS	TR-04-072225MS	PO112456.D	24 Jul 2025 11:11		YP/AJ	Ok
10	Q2674-01MSD	TR-04-072225MSD	PO112457.D	24 Jul 2025 11:30	RPD Fail	YP/AJ	Ok
11	Q2676-03	HD-02-07222025	PO112458.D	24 Jul 2025 11:48	AR1254 + 1260 Hit	YP/AJ	Ok,NS
12	Q2654-01	71125-A	PO112459.D	24 Jul 2025 12:07		YP/AJ	Ok
13	Q2658-02	PAD-072125-2	PO112460.D	24 Jul 2025 12:24		YP/AJ	Ok
14	Q2656-01	71525-A	PO112461.D	24 Jul 2025 12:43	AR1248 Hit	YP/AJ	Ok
15	Q2656-03	71525-C	PO112462.D	24 Jul 2025 13:01	AR1248 Hit	YP/AJ	Ok
16	Q2663-13	HEH700H-1-1	PO112463.D	24 Jul 2025 13:19		YP/AJ	Ok
17	AR1660CCC500	AR1660CCC500	PO112464.D	24 Jul 2025 15:23	DCB Low in 1st column	YP/AJ	Ok
18	AR1242CCC500	AR1242CCC500	PO112465.D	24 Jul 2025 15:42		YP/AJ	Ok

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO072425

Review By	yogesh	Review On	7/24/2025 11:32:48 AM
Supervise By		Supervise On	
SubDirectory	PO072425	HP Acquire Method	HP Processing Method PO070825
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP 24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	AR1248CCC500	AR1248CCC500	PO112466.D	24 Jul 2025 16:00	DCB Low in 1st column	YP/AJ	Ok
20	AR1254CCC500	AR1254CCC500	PO112467.D	24 Jul 2025 16:17		YP/AJ	Ok
21	I.BLK	I.BLK	PO112468.D	24 Jul 2025 16:36		YP/AJ	Ok
22	PB168989BL	PB168989BL	PO112469.D	24 Jul 2025 16:54		YP/AJ	Ok
23	PB168989BS	PB168989BS	PO112470.D	24 Jul 2025 17:13		YP/AJ	Ok
24	PB168990BL	PB168990BL	PO112471.D	24 Jul 2025 17:31		YP/AJ	Ok
25	PB168990BS	PB168990BS	PO112472.D	24 Jul 2025 17:49		YP/AJ	Ok
26	Q2673-01	TRE-25-0081	PO112473.D	24 Jul 2025 18:07		YP/AJ	Ok
27	Q2678-01	I-1361-USED-OIL	PO112474.D	24 Jul 2025 18:25		YP/AJ	Ok,NS
28	Q2679-01	M2239	PO112475.D	24 Jul 2025 18:44	TCMX low in 1st column	YP/AJ	Ok,NS
29	Q2688-09	SOM-81	PO112476.D	24 Jul 2025 19:01	AR1260 Hit	YP/AJ	Ok
30	Q2688-11	SOM-77	PO112477.D	24 Jul 2025 19:20		YP/AJ	Ok,NS
31	Q2689-01	OR-03-07232025	PO112478.D	24 Jul 2025 19:37		YP/AJ	Ok
32	AR1660CCC500	AR1660CCC500	PO112479.D	24 Jul 2025 20:46		YP/AJ	Ok
33	AR1242CCC500	AR1242CCC500	PO112480.D	24 Jul 2025 21:42		YP/AJ	Ok
34	AR1248CCC500	AR1248CCC500	PO112481.D	24 Jul 2025 21:59		YP/AJ	Ok
35	AR1254CCC500	AR1254CCC500	PO112482.D	24 Jul 2025 22:17		YP/AJ	Ok
36	I.BLK	I.BLK	PO112483.D	25 Jul 2025 01:40		YP/AJ	Ok
37	PB168997BL	PB168997BL	PO112484.D	25 Jul 2025 01:58	Run with cleanup sample	YP/AJ	Not Ok

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO072425

Review By	yogesh	Review On	7/24/2025 11:32:48 AM
Supervise By		Supervise On	
SubDirectory	PO072425	HP Acquire Method	HP Processing Method PO070825
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP 24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

38	PB168997BS	PB168997BS	PO112485.D	25 Jul 2025 02:15		YP/AJ	Ok,NR
39	Q2681-01	NYPA-POUCH-SPENT	PO112486.D	25 Jul 2025 02:34	DCB low 2nd column	YP/AJ	Ok,NS
40	Q2684-01	PL-02-07232025	PO112487.D	25 Jul 2025 02:52		YP/AJ	Ok,NS
41	Q2684-01MS	PL-02-07232025MS	PO112488.D	25 Jul 2025 03:11	AR1260 Recovery low	YP/AJ	Ok,NS
42	Q2684-01MSD	PL-02-07232025MSD	PO112489.D	25 Jul 2025 03:28	AR1260 Recovery low	YP/AJ	Ok,NS
43	Q2685-01	COMP-1	PO112490.D	25 Jul 2025 03:46	TCMX high in first column	YP/AJ	Ok,NS
44	Q2685-02	COMP-2	PO112491.D	25 Jul 2025 04:05		YP/AJ	Ok,NS
45	Q2685-03	COMP-3	PO112492.D	25 Jul 2025 04:23		YP/AJ	Ok,NS
46	Q2685-04	COMP-4	PO112493.D	25 Jul 2025 04:40	AR1260 hits	YP/AJ	Ok
47	AR1660CCC500	AR1660CCC500	PO112494.D	25 Jul 2025 05:52	DCB low in 1st column,AR1260-01,02,05 low in 1st column	YP/AJ	Ok,NS
48	AR1242CCC500	AR1242CCC500	PO112495.D	25 Jul 2025 07:31		YP/AJ	Ok
49	AR1254CCC500	AR1254CCC500	PO112496.D	25 Jul 2025 07:50	TCMX high in first column ,DCB low in first column ,AR1242-05 low in first column	YP/AJ	Ok,NS
50	I.BLK	I.BLK	PO112497.D	25 Jul 2025 08:08		YP/AJ	Ok

M : Manual Integration

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO072525

Review By	yogesh	Review On	7/25/2025 12:46:20 PM
Supervise By		Supervise On	
SubDirectory	PO072525	HP Acquire Method	HP Processing Method PO070825

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369
CCC	PP24332,PP24347,PP24352,PP24357
Internal Standard/PEM	
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PO112498.D	25 Jul 2025 09:06		YP/AJ	Ok
2	AR1660CCC500	AR1660CCC500	PO112499.D	25 Jul 2025 09:45	AR1260 - 4,5 low in 1st column	YP/AJ	Ok,NS
3	AR1242CCC500	AR1242CCC500	PO112500.D	25 Jul 2025 10:05		YP/AJ	Ok
4	AR1248CCC500	AR1248CCC500	PO112501.D	25 Jul 2025 10:24		YP/AJ	Ok
5	AR1254CCC500	AR1254CCC500	PO112502.D	25 Jul 2025 10:43		YP/AJ	Ok
6	I.BLK	I.BLK	PO112503.D	25 Jul 2025 11:01		YP/AJ	Ok
7	DDT ANALOG	DDT ANALOG	PO112504.D	25 Jul 2025 11:20		YP/AJ	Ok
8	Q2687-01	VNJ-251	PO112505.D	25 Jul 2025 11:57		YP/AJ	Ok
9	Q2688-01	COMP-1	PO112506.D	25 Jul 2025 12:14		YP/AJ	Ok,NS
10	Q2688-02	COMP-2	PO112507.D	25 Jul 2025 12:32		YP/AJ	Ok,NS
11	Q2688-03	COMP-3	PO112508.D	25 Jul 2025 12:50		YP/AJ	Ok,NS
12	Q2688-04	COMP-4	PO112509.D	25 Jul 2025 13:08		YP/AJ	Ok,NS
13	PB168997BL	PB168997BL	PO112510.D	25 Jul 2025 13:26		YP/AJ	Ok
14	Q2687-01	VNJ-251	PO112511.D	25 Jul 2025 13:43		YP/AJ	Ok
15	Q2683-03	DSN001	PO112512.D	25 Jul 2025 14:00	TCMX low in both column	YP/AJ	ReRun
16	Q2683-07	DSN003	PO112513.D	25 Jul 2025 14:19	TCMX low in both column	YP/AJ	ReRun
17	AR1660CCC500	AR1660CCC500	PO112514.D	25 Jul 2025 15:32		YP/AJ	Ok,NS
18	AR1242CCC500	AR1242CCC500	PO112515.D	25 Jul 2025 15:49		YP/AJ	Ok

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO072525

Review By	yogesh	Review On	7/25/2025 12:46:20 PM
Supervise By		Supervise On	
SubDirectory	PO072525	HP Acquire Method	HP Processing Method PO070825
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP 24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	AR1248CCC500	AR1248CCC500	PO112516.D	25 Jul 2025 16:06		YP/AJ	Ok
20	AR1254CCC500	AR1254CCC500	PO112517.D	25 Jul 2025 16:25		YP/AJ	Ok
21	I.BLK	I.BLK	PO112518.D	25 Jul 2025 16:43		YP/AJ	Ok
22	PB169007BL	PB169007BL	PO112519.D	25 Jul 2025 17:02		YP/AJ	Ok
23	PB169007BS	PB169007BS	PO112520.D	25 Jul 2025 17:20		YP/AJ	Ok,NS
24	PB169007BSD	PB169007BSD	PO112521.D	25 Jul 2025 17:38		YP/AJ	Ok,NS
25	PB169006BL	PB169006BL	PO112522.D	25 Jul 2025 17:55	Already run	YP/AJ	Not Ok
26	PB169006BS	PB169006BS	PO112523.D	25 Jul 2025 18:13		YP/AJ	Ok
27	PB169006BSD	PB169006BSD	PO112524.D	25 Jul 2025 18:30		YP/AJ	Ok
28	Q2688-12	SOM-78	PO112525.D	25 Jul 2025 18:47	DCB Low in both column ,AR1260 Hit	YP/AJ	ReRun
29	Q2688-13	TOTE-COMP	PO112526.D	25 Jul 2025 19:05	Surrogate fail	YP/AJ	Not Ok
30	Q2672-05	AUD-25-0122	PO112527.D	25 Jul 2025 19:23		YP/AJ	Ok,NS
31	Q2691-05	298	PO112528.D	25 Jul 2025 19:41	Both surrogate high in both column	YP/AJ	Not Ok
32	AR1660CCC500	AR1660CCC500	PO112529.D	25 Jul 2025 20:53		YP/AJ	Ok,NS
33	AR1242CCC500	AR1242CCC500	PO112530.D	25 Jul 2025 21:49		YP/AJ	Ok,NS
34	AR1248CCC500	AR1248CCC500	PO112531.D	25 Jul 2025 22:07		YP/AJ	Ok
35	AR1254CCC500	AR1254CCC500	PO112532.D	25 Jul 2025 22:26		YP/AJ	Ok
36	I.BLK	I.BLK	PO112533.D	25 Jul 2025 22:44		YP/AJ	Ok

M : Manual Integration