

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** CORE02
Lab Code: ACE **SDG NO.:** Q2681
GC Column: RTX-CLP **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/29/2025 07/29/2025

Client Sample No.: CCAL01 **Date Analyzed:** 07/29/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031293.D **Time Analyzed:** 23:45

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-TP(Silvex)	9.324	9.227	9.427	716.860	712.500	0.6
2,4-D	8.444	8.346	8.546	695.390	705.000	-1.4
2,4-DCAA	7.313	7.215	7.415	750.570	750.000	0.1

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** CORE02
Lab Code: ACE **SDG NO.:** Q2681
GC Column: RTX-CLP2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/29/2025 07/29/2025

Client Sample No.: CCAL01 **Date Analyzed:** 07/29/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031293.D **Time Analyzed:** 23:45

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-TP(Silvex)	9.923	9.824	10.024	706.120	712.500	-0.9
2,4-D	9.024	8.925	9.125	687.020	705.000	-2.6
2,4-DCAA	7.765	7.666	7.866	736.100	750.000	-1.9