

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** CORE02
Lab Code: ACE **SDG NO.:** Q2681
GC Column: RTX-CLP **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/29/2025 07/29/2025

Client Sample No.: CCAL02 **Date Analyzed:** 07/30/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031301.D **Time Analyzed:** 02:57

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.322	9.227	9.427	723.660	712.500	1.6
2,4-D	8.441	8.346	8.546	708.400	705.000	0.5
2,4-DCAA	7.312	7.215	7.415	769.970	750.000	2.7

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** CORE02
Lab Code: ACE **SDG NO.:** Q2681
GC Column: RTX-CLP2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/29/2025 07/29/2025

Client Sample No.: CCAL02 **Date Analyzed:** 07/30/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031301.D **Time Analyzed:** 02:57

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.923	9.824	10.024	709.370	712.500	-0.4
2,4-D	9.023	8.925	9.125	689.660	705.000	-2.2
2,4-DCAA	7.764	7.666	7.866	735.430	750.000	-1.9