

Instrument ID: ECD\_S

**Daily Analysis Runlog For Sequence/QC Batch ID # PS072925**

|              |          |                   |                                    |
|--------------|----------|-------------------|------------------------------------|
| Review By    | Abdul    | Review On         | 7/30/2025 8:35:13 AM               |
| Supervise By |          | Supervise On      |                                    |
| SubDirectory | PS072925 | HP Acquire Method | HP Processing Method pd072925 8081 |

| STD. NAME                               | STD REF.#                                               |
|-----------------------------------------|---------------------------------------------------------|
| Tune/Reschk<br>Initial Calibration Stds | PP24553,PP24554,PP24556,PP24557,PP24558,PP24559,PP24560 |
| CCC                                     | PP24559                                                 |
| Internal Standard/PEM                   |                                                         |
| ICV/I.BLK                               | PP24562                                                 |
| Surrogate Standard                      |                                                         |
| MS/MSD Standard                         |                                                         |
| LCS Standard                            |                                                         |

| Sr# | SampleId    | ClientID    | Data File Name | Date-Time         | Comment                       | Operator | Status |
|-----|-------------|-------------|----------------|-------------------|-------------------------------|----------|--------|
| 1   | HEXANE      | HEXANE      | PS031273.D     | 29 Jul 2025 15:42 |                               | AR\AJ    | Ok     |
| 2   | I.BLK       | I.BLK       | PS031274.D     | 29 Jul 2025 16:06 |                               | AR\AJ    | Ok     |
| 3   | HSTDICC200  | HSTDICC200  | PS031275.D     | 29 Jul 2025 16:30 |                               | AR\AJ    | Ok,NS  |
| 4   | HSTDICC500  | HSTDICC500  | PS031276.D     | 29 Jul 2025 16:54 |                               | AR\AJ    | Ok     |
| 5   | HSTDICC750  | HSTDICC750  | PS031277.D     | 29 Jul 2025 17:18 |                               | AR\AJ    | Ok     |
| 6   | HSTDICC1000 | HSTDICC1000 | PS031278.D     | 29 Jul 2025 17:42 |                               | AR\AJ    | Ok,NS  |
| 7   | HSTDICC1500 | HSTDICC1500 | PS031279.D     | 29 Jul 2025 18:07 |                               | AR\AJ    | Ok,NS  |
| 8   | HSTDICV750  | ICVPS072925 | PS031280.D     | 29 Jul 2025 18:31 |                               | AR\AJ    | Ok     |
| 9   | I.BLK       | I.BLK       | PS031281.D     | 29 Jul 2025 18:55 |                               | AR\AJ    | Ok     |
| 10  | HSTDCCC750  | HSTDCCC750  | PS031282.D     | 29 Jul 2025 19:19 |                               | AR\AJ    | Ok,NS  |
| 11  | Q2668-01    | TP-2        | PS031283.D     | 29 Jul 2025 19:43 |                               | AR\AJ    | Ok     |
| 12  | Q2668-05    | TP-3        | PS031284.D     | 29 Jul 2025 20:07 |                               | AR\AJ    | Ok     |
| 13  | Q2668-09    | TP-1        | PS031285.D     | 29 Jul 2025 20:32 |                               | AR\AJ    | Ok     |
| 14  | PB169024BL  | PB169024BL  | PS031286.D     | 29 Jul 2025 20:56 |                               | AR\AJ    | Ok     |
| 15  | PB169024BS  | PB169024BS  | PS031287.D     | 29 Jul 2025 21:20 |                               | AR\AJ    | Ok,NS  |
| 16  | Q2668-09MS  | TP-1MS      | PS031288.D     | 29 Jul 2025 21:44 | Some compound recovery fail   | AR\AJ    | Ok,NS  |
| 17  | Q2668-09MSD | TP-1MSD     | PS031289.D     | 29 Jul 2025 22:08 | Some compound recovery fail   | AR\AJ    | Ok,NS  |
| 18  | Q2691-01    | 295         | PS031290.D     | 29 Jul 2025 22:32 | surrogate fail in both column | AR\AJ    | Not Ok |

Instrument ID: ECD\_S

**Daily Analysis Runlog For Sequence/QC Batch ID # PS072925**

| Review By                                                          | Abdul                                                   | Review On         | 7/30/2025 8:35:13 AM |
|--------------------------------------------------------------------|---------------------------------------------------------|-------------------|----------------------|
| Supervise By                                                       |                                                         | Supervise On      |                      |
| SubDirectory                                                       | PS072925                                                | HP Acquire Method | HP Processing Method |
|                                                                    |                                                         |                   | pd072925 8081        |
| STD. NAME                                                          | STD REF.#                                               |                   |                      |
| Tune/Reschk<br>Initial Calibration Stds                            | PP24553,PP24554,PP24556,PP24557,PP24558,PP24559,PP24560 |                   |                      |
| CCC<br>Internal Standard/PEM                                       | PP24559                                                 |                   |                      |
| ICV/I.BLK<br>Surrogate Standard<br>MS/MSD Standard<br>LCS Standard | PP24562                                                 |                   |                      |

|    |             |                  |            |                   |                           |       |       |
|----|-------------|------------------|------------|-------------------|---------------------------|-------|-------|
| 19 | Q2691-07    | 299              | PS031291.D | 29 Jul 2025 22:57 | Surrogate fail in 1st col | AR\AJ | ReRun |
| 20 | I.BLK       | I.BLK            | PS031292.D | 29 Jul 2025 23:21 |                           | AR\AJ | Ok    |
| 21 | HSTDCCC750  | HSTDCCC750       | PS031293.D | 29 Jul 2025 23:45 |                           | AR\AJ | Ok,NS |
| 22 | PB169037BL  | PB169037BL       | PS031294.D | 30 Jul 2025 00:09 |                           | AR\AJ | Ok    |
| 23 | PB169037BS  | PB169037BS       | PS031295.D | 30 Jul 2025 00:33 |                           | AR\AJ | Ok    |
| 24 | PB168986TB  | PB168986TB       | PS031296.D | 30 Jul 2025 00:57 |                           | AR\AJ | Ok    |
| 25 | Q2681-01    | NYPA-POUCH-SPENT | PS031297.D | 30 Jul 2025 01:20 |                           | AR\AJ | Ok    |
| 26 | Q2681-01MS  | NYPA-POUCH-SPENT | PS031298.D | 30 Jul 2025 01:44 | Recovery fail             | AR\AJ | Ok,NS |
| 27 | Q2681-01MSD | NYPA-POUCH-SPENT | PS031299.D | 30 Jul 2025 02:09 | Recovery fail             | AR\AJ | Ok,NS |
| 28 | I.BLK       | I.BLK            | PS031300.D | 30 Jul 2025 02:33 |                           | AR\AJ | Ok    |
| 29 | HSTDCCC750  | HSTDCCC750       | PS031301.D | 30 Jul 2025 02:57 |                           | AR\AJ | Ok,NS |

M : Manual Integration