

### CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: CORE02

Lab Code: ACE

SDG NO.: Q2681

Continuing Calib Date: 07/29/2025

Initial Calibration Date(s): 07/28/2025

07/28/2025

Continuing Calib Time: 13:16

Initial Calibration Time(s): 16:52

17:47

GC Column: ZB-MR1 ID: 0.32 (mm)

| COMPOUND             | CCAL<br>RT | AVG<br>RT | RT WINDOW<br>FROM   TO |      | DIFF<br>RT |
|----------------------|------------|-----------|------------------------|------|------------|
| Decachlorobiphenyl   | 9.02       | 9.02      | 8.92                   | 9.12 | 0.00       |
| Tetrachloro-m-xylene | 3.53       | 3.54      | 3.44                   | 3.64 | 0.01       |
| gamma-BHC (Lindane)  | 4.31       | 4.31      | 4.21                   | 4.41 | 0.00       |
| Heptachlor           | 4.90       | 4.90      | 4.80                   | 5.00 | 0.00       |
| Heptachlor epoxide   | 5.66       | 5.66      | 5.56                   | 5.76 | 0.00       |
| Endrin               | 6.54       | 6.54      | 6.44                   | 6.64 | 0.00       |
| Methoxychlor         | 7.47       | 7.46      | 7.36                   | 7.56 | -0.01      |

### CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: CORE02

Lab Code: ACE

SDG NO.: Q2681

Continuing Calib Date: 07/29/2025

Initial Calibration Date(s): 07/28/2025

07/28/2025

Continuing Calib Time: 13:16

Initial Calibration Time(s): 16:52

17:47

GC Column: ZB-MR2 ID: 0.32 (mm)

| COMPOUND             | CCAL<br>RT | AVG<br>RT | RT WINDOW<br>FROM   TO |      | DIFF<br>RT |
|----------------------|------------|-----------|------------------------|------|------------|
| Decachlorobiphenyl   | 7.99       | 7.99      | 7.89                   | 8.09 | 0.00       |
| Tetrachloro-m-xylene | 2.83       | 2.83      | 2.73                   | 2.93 | 0.00       |
| gamma-BHC (Lindane)  | 3.66       | 3.67      | 3.57                   | 3.77 | 0.01       |
| Heptachlor           | 4.01       | 4.01      | 3.91                   | 4.11 | 0.00       |
| Heptachlor epoxide   | 4.80       | 4.80      | 4.70                   | 4.90 | 0.00       |
| Endrin               | 5.71       | 5.71      | 5.61                   | 5.81 | 0.00       |
| Methoxychlor         | 6.68       | 6.68      | 6.58                   | 6.78 | 0.00       |