

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** CORE02
Lab Code: ACE **SDG NO.:** Q2681
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/28/2025 07/28/2025

Client Sample No.: CCAL01 **Date Analyzed:** 07/29/2025
Lab Sample No.: PSTDCCC050 **Data File :** PL096626.D **Time Analyzed:** 17:17

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	9.014	8.916	9.116	51.220	50.000	2.4
Endrin	6.542	6.442	6.642	49.510	50.000	-1.0
gamma-BHC (Lindane)	4.310	4.210	4.410	51.420	50.000	2.8
Heptachlor	4.902	4.803	5.003	52.100	50.000	4.2
Heptachlor epoxide	5.662	5.562	5.762	51.800	50.000	3.6
Methoxychlor	7.461	7.362	7.562	50.320	50.000	0.6
Tetrachloro-m-xylene	3.535	3.435	3.635	51.480	50.000	3.0

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** CORE02
Lab Code: ACE **SDG NO.:** Q2681
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/28/2025 07/28/2025

Client Sample No.: CCAL01 **Date Analyzed:** 07/29/2025
Lab Sample No.: PSTDCCC050 **Data File :** PL096626.D **Time Analyzed:** 17:17

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	7.991	7.892	8.092	50.050	50.000	0.1
Endrin	5.709	5.609	5.809	53.290	50.000	6.6
gamma-BHC (Lindane)	3.665	3.565	3.765	54.670	50.000	9.3
Heptachlor	4.014	3.914	4.114	53.830	50.000	7.7
Heptachlor epoxide	4.799	4.699	4.899	53.440	50.000	6.9
Methoxychlor	6.680	6.581	6.781	50.170	50.000	0.3
Tetrachloro-m-xylene	2.828	2.728	2.928	54.190	50.000	8.4