

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** CORE02
Lab Code: ACE **SDG NO.:** Q2681
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/28/2025 07/28/2025

Client Sample No.: CCAL02 **Date Analyzed:** 07/29/2025
Lab Sample No.: PSTDCCC050 **Data File :** PL096634.D **Time Analyzed:** 19:06

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	9.015	8.916	9.116	50.570	50.000	1.1
Endrin	6.542	6.442	6.642	50.490	50.000	1.0
gamma-BHC (Lindane)	4.311	4.210	4.410	50.850	50.000	1.7
Heptachlor	4.903	4.803	5.003	51.800	50.000	3.6
Heptachlor epoxide	5.662	5.562	5.762	51.380	50.000	2.8
Methoxychlor	7.462	7.362	7.562	49.820	50.000	-0.4
Tetrachloro-m-xylene	3.535	3.435	3.635	50.770	50.000	1.5

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** CORE02
Lab Code: ACE **SDG NO.:** Q2681
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/28/2025 07/28/2025

Client Sample No.: CCAL02 **Date Analyzed:** 07/29/2025
Lab Sample No.: PSTDCCC050 **Data File :** PL096634.D **Time Analyzed:** 19:06

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	7.992	7.892	8.092	53.500	50.000	7.0
Endrin	5.710	5.609	5.809	53.520	50.000	7.0
gamma-BHC (Lindane)	3.665	3.565	3.765	54.040	50.000	8.1
Heptachlor	4.015	3.914	4.114	53.220	50.000	6.4
Heptachlor epoxide	4.799	4.699	4.899	53.230	50.000	6.5
Methoxychlor	6.681	6.581	6.781	51.370	50.000	2.7
Tetrachloro-m-xylene	2.828	2.728	2.928	53.200	50.000	6.4