

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.:	<u>Q2681</u>	Analytical Method:	<u>8081B</u>
Client:	<u>Core Environmental Consultants and Serv</u>	DataFile :	<u>PL096632.D</u>

	Parameter	Spike	Sample		Units	Rec	Rec		RPD		Limits	
			Result	Result			Qual	RPD	Qual	Low	High	RPD
Lab Sample ID:	Q2681-01MSD	Client Sample ID:		NYPA-POUCH-SPENT-C/								
	(Column 1)											
	gamma-BHC (Lindane)	5	0	4.90	ug/L	98		2		60	152	20
	Heptachlor	5	0	4.70	ug/L	94		2		56	147	20
	Heptachlor epoxide	5	0	5.00	ug/L	100		2		77	143	20
	Endrin	5	0	5.00	ug/L	100		0		76	144	20
	Methoxychlor	5	0	4.80	ug/L	96		0		70	142	20
Lab Sample ID:	Q2681-01MSD	Client Sample ID:		NYPA-POUCH-SPENT-C/								
	(Column 2)											
	gamma-BHC (Lindane)	5	0	5.10	ug/L	102		2		60	152	20
	Heptachlor	5	0	4.80	ug/L	96		2		56	147	20
	Heptachlor epoxide	5	0	5.10	ug/L	102		2		77	143	20
	Endrin	5	0	5.20	ug/L	104		4		76	144	20
	Methoxychlor	5	0	5.00	ug/L	100		2		70	142	20