

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2681

Analytical Method: 8081B

Client: Core Environmental Consultants and Serv

Datafile : PL096628.D

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD		Limits	
								Qual	Low	High	RPD
PB169038BS (Column 1)	gamma-BHC (Lindane)	0.5	0.48	ug/L	97				82	129	
	Heptachlor	0.5	0.49	ug/L	98				79	127	
	Heptachlor epoxide	0.5	0.49	ug/L	98				81	124	
	Endrin	0.5	0.47	ug/L	94				81	128	
	Methoxychlor	0.5	0.46	ug/L	93				78	108	
PB169038BS (Column 2)	gamma-BHC (Lindane)	0.5	0.51	ug/L	102				82	129	
	Heptachlor	0.5	0.50	ug/L	100				79	127	
	Heptachlor epoxide	0.5	0.50	ug/L	101				81	124	
	Endrin	0.5	0.51	ug/L	103				81	128	
	Methoxychlor	0.5	0.48	ug/L	95				78	108	