

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2681

Analytical Method: 8270E

Client: Core Environmental Consultants and Services, Inc.

DataFile: BP025263.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD		Limits	
								Qual	Low	High	RPD
PB169045BS	Pyridine	50	35.2	ug/L	70				29	97	
	1,4-Dichlorobenzene	50	42.8	ug/L	86				76	103	
	2-Methylphenol	50	46.9	ug/L	94				69	109	
	3+4-Methylphenols	50	46.2	ug/L	92				67	106	
	Hexachloroethane	50	42.4	ug/L	85				76	118	
	Nitrobenzene	50	45.3	ug/L	91				58	106	
	Hexachlorobutadiene	50	41.9	ug/L	84				69	101	
	2,4,6-Trichlorophenol	50	44.8	ug/L	90				61	110	
	2,4,5-Trichlorophenol	50	44.6	ug/L	89				70	106	
	2,4-Dinitrotoluene	50	48.1	ug/L	96				60	115	
	Hexachlorobenzene	50	44.5	ug/L	89				73	106	
	Pentachlorophenol	100	81.7	ug/L	82				47	114	