

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX072125

Review By	John Carlone	Review On	7/22/2025 8:51:02 AM
Supervise By	Maresh Dadoda	Supervise On	7/22/2025 1:41:20 PM
SubDirectory	VX072125	HP Acquire Method	HP Processing Method 82X072125W.M

STD. NAME	STD REF.#
Tune/Reschk	VP134843
Initial Calibration Stds	VP134844,VP134845,VP134846,VP134847,VP134848,VP134849
CCC	
Internal Standard/PEM	
ICV/I.BLK	VP134850
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX047054.D	21 Jul 2025 08:53		JC/MD	Ok
2	VSTDICC001	VSTDICC001	VX047055.D	21 Jul 2025 09:47		JC/MD	Ok,M
3	VSTDICC005	VSTDICC005	VX047056.D	21 Jul 2025 10:08	Comp. #11 is on Linear Regression	JC/MD	Ok,M
4	VSTDICC020	VSTDICC020	VX047057.D	21 Jul 2025 10:29		JC/MD	Ok,M
5	VSTDICCC050	VSTDICCC050	VX047058.D	21 Jul 2025 10:50		JC/MD	Ok,M
6	VSTDICC100	VSTDICC100	VX047059.D	21 Jul 2025 11:11		JC/MD	Ok,M
7	VSTDICC150	VSTDICC150	VX047060.D	21 Jul 2025 11:32		JC/MD	Ok,M
8	IBLK	IBLK	VX047061.D	21 Jul 2025 11:52		JC/MD	Ok
9	VSTDICV050	ICVVX072125	VX047062.D	21 Jul 2025 13:17		JC/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX072825

Review By	John Carlone	Review On	7/29/2025 8:49:52 AM
Supervise By	Maresh Dadoda	Supervise On	7/29/2025 12:09:18 PM
SubDirectory	VX072825	HP Acquire Method	HP Processing Method 82X072125W.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP134919
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP134920,VP134921

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX047151.D	28 Jul 2025 08:52		JC/MD	Ok
2	VSTDCCC050	VSTDCCC050	VX047152.D	28 Jul 2025 09:23	pH#Lot#V12668	JC/MD	Ok,M
3	VX0728MBL01	VX0728MBL01	VX047153.D	28 Jul 2025 09:58		JC/MD	Ok
4	VX0728WBL01	VX0728WBL01	VX047154.D	28 Jul 2025 10:19		JC/MD	Ok
5	VX0728WBS01	VX0728WBS01	VX047155.D	28 Jul 2025 10:45		JC/MD	Ok,M
6	VX0728WBSD01	VX0728WBSD01	VX047156.D	28 Jul 2025 11:11		JC/MD	Ok,M
7	IBLK	IBLK	VX047157.D	28 Jul 2025 11:32		JC/MD	Ok
8	Q2693-01	STORAGE-BLANK-SO	VX047158.D	28 Jul 2025 11:53	vial B pH<2	JC/MD	Ok
9	Q2693-02	STORAGE-BLANK-WA	VX047159.D	28 Jul 2025 12:20	vial B pH<2	JC/MD	Ok
10	Q2693-03	STORAGE-BLANK-WA	VX047160.D	28 Jul 2025 13:16	vial B pH<2	JC/MD	Ok
11	PB169005TB	PB169005TB	VX047161.D	28 Jul 2025 13:42		JC/MD	Ok
12	Q2681-01	NYPA-POUCH-SPENT	VX047162.D	28 Jul 2025 14:03	vial A pH#5.0 Surrogate fail;Need 5X	JC/MD	Dilution
13	Q2691-02	295	VX047163.D	28 Jul 2025 14:24	vial A pH#5.0	JC/MD	Ok
14	Q2691-08	299	VX047164.D	28 Jul 2025 14:46	vial A pH#5.0	JC/MD	Ok,M
15	Q2681-01DL	NYPA-POUCH-SPENT	VX047165.D	28 Jul 2025 15:07	vial B pH#5.0	JC/MD	Ok
16	Q2706-02	RT-5417	VX047166.D	28 Jul 2025 15:28	vial A pH#5.0	JC/MD	Ok,M
17	Q2706-04	ETGI-361	VX047167.D	28 Jul 2025 15:50	vial A pH#5.0	JC/MD	Ok,M

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX072825

Review By	John Carlone	Review On	7/29/2025 8:49:52 AM
Supervise By	Maresh Dadoda	Supervise On	7/29/2025 12:09:18 PM
SubDirectory	VX072825	HP Acquire Method	HP Processing Method 82X072125W.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP134919
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP134920,VP134921

18	Q2703-04	TP-4	VX047168.D	28 Jul 2025 16:11	vial A pH#5.0	JC/MD	Ok,M
19	VSTDCCC050	VSTDCCC050EC	VX047169.D	28 Jul 2025 16:39		JC/MD	Ok,M

M : Manual Integration