

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q2681</u>	Analytical Method:	<u>SW8260-Low</u>
Client:	<u>Core Environmental Consultants and</u>	Datafile :	<u>VX047155.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX0728WBS01	Vinyl chloride	20	19.3	ug/L	97			65	117	
	1,1-Dichloroethene	20	18.4	ug/L	92			74	110	
	2-Butanone	100	92.4	ug/L	92			65	122	
	Carbon Tetrachloride	20	18.9	ug/L	95			77	113	
	Chloroform	20	18.5	ug/L	93			79	113	
	Benzene	20	18.7	ug/L	94			82	109	
	1,2-Dichloroethane	20	19.2	ug/L	96			80	115	
	Trichloroethene	20	18.5	ug/L	93			77	113	
	Tetrachloroethene	20	18.3	ug/L	92			67	123	
	Chlorobenzene	20	18.6	ug/L	93			82	109	