

Laboratory Control Sample Summary

Client: Parkway Generation Operating LLC

SDG No.: Q2683

Project: Linden Generating Station - LRSA

Run No.: LB136596

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136596BS							
BOD5	mg/L	198	191		96	1	84.6-115.4	07/24/2025

Laboratory Control Sample Summary

Client: Parkway Generation Operating LLC

SDG No.: Q2683

Project: Linden Generating Station - LRSA

Run No.: LB136600

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136600BS							
TDS	mg/L	100	95.0		95	1	90-110	07/24/2025

Laboratory Control Sample Summary

Client: Parkway Generation Operating LLC

SDG No.: Q2683

Project: Linden Generating Station - LRSA

Run No.: LB136616

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136616BS							
COD	mg/L	50	47.9		96	1	90-110	07/25/2025

Laboratory Control Sample Summary

Client: Parkway Generation Operating LLC

SDG No.: Q2683

Project: Linden Generating Station - LRSA

Run No.: LB136626

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136626BS							
TSS	mg/L	550	532		97	1	90-110	07/25/2025

Laboratory Control Sample Summary

Client: Parkway Generation Operating LLC

SDG No.: Q2683

Project: Linden Generating Station - LRSA

Run No.: LB136615

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB169011BS							
Ammonia as N	mg/L	1	1.00		100	1	90-110	07/25/2025