

Instrument ID: **MSVOA_X**

Daily Analysis Runlog For Sequence/QC Batch ID # VX072125

Review By	John Carlone	Review On	7/22/2025 8:51:02 AM
Supervise By	Maresh Dadoda	Supervise On	7/22/2025 1:41:20 PM
SubDirectory	VX072125	HP Acquire Method	HP Processing Method 82X072125W.M
STD. NAME	STD REF.#		
Tune/Reschk	VP134843		
Initial Calibration Stds	VP134844,VP134845,VP134846,VP134847,VP134848,VP134849		
CCC			
Internal Standard/PEM			
ICV/I.BLK	VP134850		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	BFB	VX047054.D	21 Jul 2025 08:53	JC/MD	Ok
2	VSTDIC001	VX047055.D	21 Jul 2025 09:47	JC/MD	Ok,M
3	VSTDIC005	VX047056.D	21 Jul 2025 10:08	JC/MD	Ok,M
4	VSTDIC020	VX047057.D	21 Jul 2025 10:29	JC/MD	Ok,M
5	VSTDIC050	VX047058.D	21 Jul 2025 10:50	JC/MD	Ok,M
6	VSTDIC100	VX047059.D	21 Jul 2025 11:11	JC/MD	Ok,M
7	VSTDIC150	VX047060.D	21 Jul 2025 11:32	JC/MD	Ok,M
8	IBLK	VX047061.D	21 Jul 2025 11:52	JC/MD	Ok
9	VSTDICV050	VX047062.D	21 Jul 2025 13:17	JC/MD	Ok,M

M : Manual Integration

Instrument ID: **MSVOA_X**

Daily Analysis Runlog For Sequence/QC Batch ID # VX072525

Review By	John Carlone	Review On	7/28/2025 8:34:38 AM
Supervise By	Maresh Dadoda	Supervise On	7/28/2025 2:58:07 PM
SubDirectory	VX072525	HP Acquire Method	HP Processing Method 82X072125W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP134915,VP134918		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP134916,VP134917		

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	BFB	VX047127.D	25 Jul 2025 08:40	JC/MD	Ok
2	VSTDCCC050	VX047128.D	25 Jul 2025 09:40	JC/MD	Ok,M
3	VX0725MBL01	VX047129.D	25 Jul 2025 10:08	JC/MD	Ok
4	VX0725WBL01	VX047130.D	25 Jul 2025 10:29	JC/MD	Ok
5	VX0725MBS01	VX047131.D	25 Jul 2025 10:50	JC/MD	Ok,M
6	VX0725WBS01	VX047132.D	25 Jul 2025 11:17	JC/MD	Ok,M
7	Q2655-01MEDL	VX047133.D	25 Jul 2025 11:38	JC/MD	Not Ok
8	VX0725WBSD01	VX047134.D	25 Jul 2025 12:00	JC/MD	Ok,M
9	IBLK	VX047135.D	25 Jul 2025 12:21	JC/MD	Ok
10	Q2693-01	VX047136.D	25 Jul 2025 12:42	JC/MD	ReRun
11	Q2693-02	VX047137.D	25 Jul 2025 13:04	JC/MD	ReRun
12	Q2693-03	VX047138.D	25 Jul 2025 13:25	JC/MD	ReRun
13	Q2693-04	VX047139.D	25 Jul 2025 13:47	JC/MD	Ok
14	Q2655-01ME	VX047140.D	25 Jul 2025 14:08	JC/MD	Ok,M
15	BFB	VX047141.D	25 Jul 2025 14:30	JC/MD	Ok
16	VSTDICC005	VX047142.D	25 Jul 2025 15:18	JC/MD	Not Ok
17	VSTDICCC020	VX047143.D	25 Jul 2025 15:40	JC/MD	Ok,M
18	VSTDICC050	VX047144.D	25 Jul 2025 16:01	JC/MD	Ok,M
19	VSTDICC100	VX047145.D	25 Jul 2025 16:23	JC/MD	Ok,M
20	VSTDICC150	VX047146.D	25 Jul 2025 16:44	JC/MD	Ok,M
21	IBLK	VX047147.D	25 Jul 2025 17:06	JC/MD	Ok,M

Instrument ID: **MSVOA_X**

Daily Analysis Runlog For Sequence/QC Batch ID # VX072525

Review By	John Carlone	Review On	7/28/2025 8:34:38 AM
Supervise By	Mahesh Dadoda	Supervise On	7/28/2025 2:58:07 PM
SubDirectory	VX072525	HP Acquire Method	HP Processing Method 82X072125W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP134915,VP134918		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP134916,VP134917		

22	VSTDIC005	VX047148.D	25 Jul 2025 17:27	JC/MD	Ok,M
23	VSTDIC020	VX047149.D	25 Jul 2025 17:49	JC/MD	Not Ok
24	VSTDICV020	VX047150.D	25 Jul 2025 18:10	JC/MD	Ok,M

M : Manual Integration

Instrument ID: **MSVOA_X**

Daily Analysis Runlog For Sequence/QC Batch ID # VX072825

Review By	John Carlone	Review On	7/29/2025 8:49:52 AM
Supervise By	Maresh Dadoda	Supervise On	7/29/2025 12:09:18 PM
SubDirectory	VX072825	HP Acquire Method	HP Processing Method 82X072125W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP134919		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP134920,VP134921		

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	BFB	VX047151.D	28 Jul 2025 08:52	JC/MD	Ok
2	VSTDCCC050	VX047152.D	28 Jul 2025 09:23	JC/MD	Ok,M
3	VX0728MBL01	VX047153.D	28 Jul 2025 09:58	JC/MD	Ok
4	VX0728WBL01	VX047154.D	28 Jul 2025 10:19	JC/MD	Ok
5	VX0728WBS01	VX047155.D	28 Jul 2025 10:45	JC/MD	Ok,M
6	VX0728WBSD01	VX047156.D	28 Jul 2025 11:11	JC/MD	Ok,M
7	IBLK	VX047157.D	28 Jul 2025 11:32	JC/MD	Ok
8	Q2693-01	VX047158.D	28 Jul 2025 11:53	JC/MD	Ok
9	Q2693-02	VX047159.D	28 Jul 2025 12:20	JC/MD	Ok
10	Q2693-03	VX047160.D	28 Jul 2025 13:16	JC/MD	Ok
11	PB169005TB	VX047161.D	28 Jul 2025 13:42	JC/MD	Ok
12	Q2681-01	VX047162.D	28 Jul 2025 14:03	JC/MD	Dilution
13	Q2691-02	VX047163.D	28 Jul 2025 14:24	JC/MD	Ok
14	Q2691-08	VX047164.D	28 Jul 2025 14:46	JC/MD	Ok,M
15	Q2681-01DL	VX047165.D	28 Jul 2025 15:07	JC/MD	Ok
16	Q2706-02	VX047166.D	28 Jul 2025 15:28	JC/MD	Ok,M
17	Q2706-04	VX047167.D	28 Jul 2025 15:50	JC/MD	Ok,M
18	Q2703-04	VX047168.D	28 Jul 2025 16:11	JC/MD	Ok,M
19	VSTDCCC050	VX047169.D	28 Jul 2025 16:39	JC/MD	Ok,M

M : Manual Integration