

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: CHEM02
 Lab Code: ACE SDG NO.: Q2693
 Lab File ID: VX047128.D Date Analyzed: 07/25/2025
 Instrument ID: MSVOA_X Time Analyzed: 09:40
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	277253	5.56	466924	6.76	414717	10.05
UPPER LIMIT	554506	6.056	933848	7.263	829434	10.549
LOWER LIMIT	138627	5.056	233462	6.263	207359	9.549
EPA SAMPLE NO.						
STORAGE-BLANK-SAMPLE-MANAGEMENT	397105	5.57	725707	6.77	694329	10.06
VX0725WBL01	437780	5.56	806824	6.77	773720	10.06
VX0725WBS01	280628	5.56	481726	6.76	427937	10.06
VX0725WBSD01	296074	5.56	508838	6.77	457935	10.06

IS1 = Pentafluorobenzene

IS2 = 1,4-Difluorobenzene

IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.