

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: Alliance

Contract: CHEM02

Lab Code: ACE

SDG NO.: Q2693

Lab File ID: VX047151.D

BFB Injection Date: 07/28/2025

Instrument ID: MSVOA_X

BFB Injection Time: 08:52

GC Column: DB-624UI ID: 0.18 (mm)

Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	19.3
75	30.0 - 60.0% of mass 95	52
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.4
173	Less than 2.0% of mass 174	0.7 (0.9) 1
174	50.0 - 100.0% of mass 95	72.3
175	5.0 - 9.0% of mass 174	5.4 (7.4) 1
176	95.0 - 101.0% of mass 174	70 (96.8) 1
177	5.0 - 9.0% of mass 176	4.6 (6.5) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

CLIENT ID	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC050	VSTDCCC050	VX047152.D	07/28/2025	09:23
VX0728WBL01	VX0728WBL01	VX047154.D	07/28/2025	10:19
VX0728WBS01	VX0728WBS01	VX047155.D	07/28/2025	10:45
STORAGE-BLANK-SOIL-REF-6	Q2693-01	VX047158.D	07/28/2025	11:53
STORAGE-BLANK-WATER-REF-5	Q2693-02	VX047159.D	07/28/2025	12:20
STORAGE-BLANK-WATER-REF-4	Q2693-03	VX047160.D	07/28/2025	13:16