

LAB CHRONICLE

OrderID:	Q2695	OrderDate:	7/25/2025 10:40:00 AM
Client:	Tetra Tech NUS, Inc.	Project:	NWIRP Bethpage 112G08005-WE13
Contact:	Ernie Wu	Location:	D31

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q2695-01	RW5-SP100-2025072 4	WATER	Anions Group4	300.0	07/24/25 10:45		07/25/25 13:16	07/25/25



SAMPLE DATA

A

B

C

D

Report of Analysis

Client:	Tetra Tech NUS, Inc.	Date Collected:	07/24/25 10:45
Project:	NWIRP Bethpage 112G08005-WE13	Date Received:	07/25/25
Client Sample ID:	RW5-SP100-20250724	SDG No.:	Q2695
Lab Sample ID:	Q2695-01	Matrix:	WATER
		% Solid:	0

Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
Nitrite	0.30	U	1	0.074	0.30	0.60	mg/L		07/25/25 13:16	300.0
Nitrate	3.90		1	0.095	0.25	0.50	mg/L		07/25/25 13:16	300.0
Nitrate+Nitrite	3.90		1	0.17	0.55	1.10	mg/L		07/25/25 13:16	300.0

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

D = Dilution

Q = indicates LCS control criteria did not meet requirements

H = Sample Analysis Out Of Hold Time

J = Estimated Value

B = Analyte Found in Associated Method Blank

* = indicates the duplicate analysis is not within control limits.

E = Indicates the reported value is estimated because of the presence of interference.

OR = Over Range

N =Spiked sample recovery not within control limits

QC RESULT SUMMARY

Initial and Continuing Calibration Verification

Client: Tetra Tech NUS, Inc.

SDG No.: Q2695

Project: NWIRP Bethpage 112G08005-WE13

RunNo.: LB136635

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1						
Bromide	mg/L	9.5	10	95	90-110	07/21/2025
Chloride	mg/L	2.8	3	93	90-110	07/21/2025
Fluoride	mg/L	1.9	2	95	90-110	07/21/2025
Nitrite	mg/L	2.8	3	93	90-110	07/21/2025
Nitrate	mg/L	2.3	2.5	92	90-110	07/21/2025
Sulfate	mg/L	14.2	15	95	90-110	07/21/2025
Orthophosphate as P	mg/L	4.8	5	96	90-110	07/21/2025
Sample ID: CCV1						
Bromide	mg/L	10.1	10	101	90-110	07/25/2025
Chloride	mg/L	3.1	3	103	90-110	07/25/2025
Fluoride	mg/L	2	2	100	90-110	07/25/2025
Nitrite	mg/L	3	3	100	90-110	07/25/2025
Nitrate	mg/L	2.5	2.5	100	90-110	07/25/2025
Sulfate	mg/L	14.9	15	99	90-110	07/25/2025
Orthophosphate as P	mg/L	5	5	100	90-110	07/25/2025
Sample ID: CCV2						
Bromide	mg/L	10.2	10	102	90-110	07/25/2025
Chloride	mg/L	3.1	3	103	90-110	07/25/2025
Fluoride	mg/L	2	2	100	90-110	07/25/2025
Nitrite	mg/L	3	3	100	90-110	07/25/2025
Nitrate	mg/L	2.5	2.5	100	90-110	07/25/2025
Sulfate	mg/L	15.1	15	101	90-110	07/25/2025
Orthophosphate as P	mg/L	5.1	5	102	90-110	07/25/2025
Sample ID: CCV3						
Bromide	mg/L	10.2	10	102	90-110	07/25/2025
Chloride	mg/L	3.1	3	103	90-110	07/25/2025
Fluoride	mg/L	2	2	100	90-110	07/25/2025
Nitrite	mg/L	3	3	100	90-110	07/25/2025
Nitrate	mg/L	2.5	2.5	100	90-110	07/25/2025
Sulfate	mg/L	15	15	100	90-110	07/25/2025
Orthophosphate as P	mg/L	5.1	5	102	90-110	07/25/2025

Initial and Continuing Calibration Blank Summary

Client: Tetra Tech NUS, Inc.

SDG No.: Q2695

Project: NWIRP Bethpage 112G08005-WE13

RunNo.: LB136635

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB1							
Bromide	mg/L	< 1.0000	1.0000	U	0.37	2	07/21/2025
Chloride	mg/L	< 0.3000	0.3000	U	0.19	0.6	07/21/2025
Fluoride	mg/L	< 0.2000	0.2000	U	0.11	0.4	07/21/2025
Nitrite	mg/L	< 0.3000	0.3000	U	0.074	0.6	07/21/2025
Nitrate	mg/L	< 0.2500	0.2500	U	0.095	0.5	07/21/2025
Sulfate	mg/L	< 1.5000	1.5000	U	0.46	3	07/21/2025
Orthophosphate as P	mg/L	< 0.5000	0.5000	U	0.34	1	07/21/2025
Sample ID: CCB1							
Bromide	mg/L	< 1.0000	1.0000	U	0.37	2	07/25/2025
Chloride	mg/L	< 0.3000	0.3000	U	0.19	0.6	07/25/2025
Fluoride	mg/L	< 0.2000	0.2000	U	0.11	0.4	07/25/2025
Nitrite	mg/L	< 0.3000	0.3000	U	0.074	0.6	07/25/2025
Nitrate	mg/L	< 0.2500	0.2500	U	0.095	0.5	07/25/2025
Sulfate	mg/L	< 1.5000	1.5000	U	0.46	3	07/25/2025
Orthophosphate as P	mg/L	< 0.5000	0.5000	U	0.34	1	07/25/2025
Sample ID: CCB2							
Bromide	mg/L	< 1.0000	1.0000	U	0.37	2	07/25/2025
Chloride	mg/L	< 0.3000	0.3000	U	0.19	0.6	07/25/2025
Fluoride	mg/L	< 0.2000	0.2000	U	0.11	0.4	07/25/2025
Nitrite	mg/L	< 0.3000	0.3000	U	0.074	0.6	07/25/2025
Nitrate	mg/L	< 0.2500	0.2500	U	0.095	0.5	07/25/2025
Sulfate	mg/L	< 1.5000	1.5000	U	0.46	3	07/25/2025
Orthophosphate as P	mg/L	< 0.5000	0.5000	U	0.34	1	07/25/2025
Sample ID: CCB3							
Bromide	mg/L	< 1.0000	1.0000	U	0.37	2	07/25/2025
Chloride	mg/L	< 0.3000	0.3000	U	0.19	0.6	07/25/2025
Fluoride	mg/L	< 0.2000	0.2000	U	0.11	0.4	07/25/2025
Nitrite	mg/L	< 0.3000	0.3000	U	0.074	0.6	07/25/2025
Nitrate	mg/L	< 0.2500	0.2500	U	0.095	0.5	07/25/2025
Sulfate	mg/L	< 1.5000	1.5000	U	0.46	3	07/25/2025
Orthophosphate as P	mg/L	< 0.5000	0.5000	U	0.34	1	07/25/2025

Preparation Blank Summary

Client: Tetra Tech NUS, Inc.

SDG No.: Q2695

Project: NWIRP Bethpage 112G08005-WE13

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: LB136635BLW							
Bromide	mg/L	< 1.0000	1.0000	U	0.37	2	07/25/2025
Chloride	mg/L	< 0.3000	0.3000	U	0.19	0.6	07/25/2025
Fluoride	mg/L	< 0.2000	0.2000	U	0.11	0.4	07/25/2025
Nitrite	mg/L	< 0.3000	0.3000	U	0.074	0.6	07/25/2025
Nitrate	mg/L	< 0.2500	0.2500	U	0.095	0.5	07/25/2025
Sulfate	mg/L	< 1.5000	1.5000	U	0.46	3	07/25/2025
Orthophosphate as P	mg/L	< 0.5000	0.5000	U	0.34	1	07/25/2025

Matrix Spike Summary

Client:	Tetra Tech NUS, Inc.	SDG No.:	Q2695
Project:	NWIRP Bethpage 112G08005-WE13	Sample ID:	Q2695-01
Client ID:	RW5-SP100-20250724MS	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Bromide	mg/L	80-120	10.1		0.37	U	10	1	101		07/25/2025
Chloride	mg/L	80-120	13.4	OR	10.8	OR	3	1	87		07/25/2025
Fluoride	mg/L	80-120	2.00		0.11	U	2	1	100		07/25/2025
Nitrite	mg/L	80-120	3.00		0.074	U	3	1	100		07/25/2025
Nitrate	mg/L	80-120	6.30	OR	3.90		2.5	1	96		07/25/2025
Sulfate	mg/L	80-120	16.9		2.60	J	15	1	95		07/25/2025
Orthophosphate as P	mg/L	80-120	5.10		0.34	U	5	1	102		07/25/2025

Matrix Spike Summary

Client:	Tetra Tech NUS, Inc.	SDG No.:	Q2695
Project:	NWIRP Bethpage 112G08005-WE13	Sample ID:	Q2695-01
Client ID:	RW5-SP100-20250724MSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Bromide	mg/L	80-120	10.0		0.37	U	10	1	100		07/25/2025
Chloride	mg/L	80-120	13.4	OR	10.8	OR	3	1	87		07/25/2025
Fluoride	mg/L	80-120	2.00		0.11	U	2	1	100		07/25/2025
Nitrite	mg/L	80-120	2.90		0.074	U	3	1	97		07/25/2025
Nitrate	mg/L	80-120	6.30	OR	3.90		2.5	1	96		07/25/2025
Sulfate	mg/L	80-120	16.7		2.60	J	15	1	94		07/25/2025
Orthophosphate as P	mg/L	80-120	5.00		0.34	U	5	1	100		07/25/2025

Duplicate Sample Summary

Client:	Tetra Tech NUS, Inc.	SDG No.:	Q2695
Project:	NWIRP Bethpage 112G08005-WE13	Sample ID:	Q2695-01
Client ID:	RW5-SP100-20250724MSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/AD	Qual	Analysis Date
Fluoride	mg/L	+/-20	2.00		2.00		1	0		07/25/2025
Chloride	mg/L	+/-20	13.4	OR	13.4	OR	1	0		07/25/2025
Nitrate	mg/L	+/-20	6.30	OR	6.30	OR	1	0		07/25/2025
Bromide	mg/L	+/-20	10.1		10.0		1	1		07/25/2025
Sulfate	mg/L	+/-20	16.9		16.7		1	1		07/25/2025
Orthophosphate as P	mg/L	+/-20	5.10		5.00		1	2		07/25/2025
Nitrite	mg/L	+/-20	3.00		2.90		1	3		07/25/2025

Laboratory Control Sample Summary

Client:	Tetra Tech NUS, Inc.	SDG No.:	Q2695
Project:	NWIRP Bethpage 112G08005-WE13	Run No.:	LB136635

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136635BSW							
Bromide	mg/L	10	10.2		102	1	90-110	07/25/2025
Chloride	mg/L	3	3.10		103	1	90-110	07/25/2025
Fluoride	mg/L	2	2.00		100	1	90-110	07/25/2025
Nitrite	mg/L	3	3.00		100	1	90-110	07/25/2025
Nitrate	mg/L	2.5	2.50		100	1	90-110	07/25/2025
Sulfate	mg/L	15	15.0		100	1	90-110	07/25/2025
Orthophosphate as P	mg/L	5	5.10		102	1	90-110	07/25/2025