

Initial and Continuing Calibration Verification

Client: Tetra Tech NUS, Inc.

SDG No.: Q2695

Project: NWIRP Bethpage 112G08005-WE13

RunNo.: LB136635

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1						
Bromide	mg/L	9.5	10	95	90-110	07/21/2025
Chloride	mg/L	2.8	3	93	90-110	07/21/2025
Fluoride	mg/L	1.9	2	95	90-110	07/21/2025
Nitrite	mg/L	2.8	3	93	90-110	07/21/2025
Nitrate	mg/L	2.3	2.5	92	90-110	07/21/2025
Sulfate	mg/L	14.2	15	95	90-110	07/21/2025
Orthophosphate as P	mg/L	4.8	5	96	90-110	07/21/2025
Sample ID: CCV1						
Bromide	mg/L	10.1	10	101	90-110	07/25/2025
Chloride	mg/L	3.1	3	103	90-110	07/25/2025
Fluoride	mg/L	2	2	100	90-110	07/25/2025
Nitrite	mg/L	3	3	100	90-110	07/25/2025
Nitrate	mg/L	2.5	2.5	100	90-110	07/25/2025
Sulfate	mg/L	14.9	15	99	90-110	07/25/2025
Orthophosphate as P	mg/L	5	5	100	90-110	07/25/2025
Sample ID: CCV2						
Bromide	mg/L	10.2	10	102	90-110	07/25/2025
Chloride	mg/L	3.1	3	103	90-110	07/25/2025
Fluoride	mg/L	2	2	100	90-110	07/25/2025
Nitrite	mg/L	3	3	100	90-110	07/25/2025
Nitrate	mg/L	2.5	2.5	100	90-110	07/25/2025
Sulfate	mg/L	15.1	15	101	90-110	07/25/2025
Orthophosphate as P	mg/L	5.1	5	102	90-110	07/25/2025
Sample ID: CCV3						
Bromide	mg/L	10.2	10	102	90-110	07/25/2025
Chloride	mg/L	3.1	3	103	90-110	07/25/2025
Fluoride	mg/L	2	2	100	90-110	07/25/2025
Nitrite	mg/L	3	3	100	90-110	07/25/2025
Nitrate	mg/L	2.5	2.5	100	90-110	07/25/2025
Sulfate	mg/L	15	15	100	90-110	07/25/2025
Orthophosphate as P	mg/L	5.1	5	102	90-110	07/25/2025