

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: Tetra Tech NUS, Inc.

SDG No.: Q2696

Contract: TETR06

Lab Code: ACE

ICS Source: EPA

Instrument ID: P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Iron	103000	101000	102	85600	116500	07/29/2025	13:14	LB136648
ICSAB01	Iron	98500	99300	99	84400	114500	07/29/2025	13:20	LB136648
ICSA01	Iron	99800	101000	99	85600	116500	07/30/2025	14:28	LB136658
ICSAB01	Iron	101000	99300	102	84400	114500	07/30/2025	14:33	LB136658