

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: Alliance Contract: TETR06
 Lab Code: ACE SDG No.: Q2696
 Instrument ID: BNA_N Calibration Date/Time: 07/30/2025 15:16
 Lab File ID: BN037557.D Init. Calib. Date(s): 07/15/2025 07/15/2025
 EPA Sample No.: SSTDCCC0.4EC Init. Calib. Time(s): 12:36 16:14
 GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF0.4	MIN RRF	%D	MAX%D
2-Methylnaphthalene-d10	0.574	0.502		-12.5	50.0
Fluoranthene-d10	1.060	0.915		-13.7	50.0
2-Fluorophenol	0.989	0.855		-13.5	50.0
Phenol-d6	1.241	1.046		-15.7	50.0
Nitrobenzene-d5	0.299	0.277		-7.4	50.0
2-Fluorobiphenyl	2.080	2.179		4.8	50.0
2,4,6-Tribromophenol	0.197	0.143		-27.4	50.0
Terphenyl-d14	0.859	0.818		-4.8	50.0
1,4-Dioxane	0.385	0.375		-2.6	50.0

All other compounds must meet a minimum RRF of 0.010.