

## Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2696

Analytical Method: 8270-Modified

Client: Tetra Tech NUS, Inc.

DataFile: BN037555.D

| Lab Sample ID | Parameter   | Spike | Result | Unit | Rec | RPD | Qual | RPD  | Limits |      | RPD |
|---------------|-------------|-------|--------|------|-----|-----|------|------|--------|------|-----|
|               |             |       |        |      |     |     |      | Qual | Low    | High |     |
| PB169039BS    | 1,4-Dioxane | 0.4   | 0.31   | ug/L | 78  |     |      |      | 70     | 130  |     |