

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q2696

Client ID: SSTDCCC0.4

Date Analyzed: 07/30/2025

Lab File ID: BN037548.D

Time Analyzed: 09:38

Instrument ID: BNA_N

GC Column: ZB-GR ID: 0.25 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	5852	17.087	4832	21.277	4297	23.508
UPPER LIMIT	11704	17.587	9664	21.777	8594	24.008
LOWER LIMIT	2926	16.587	2416	20.777	2148.5	23.008
EPA SAMPLE NO.						
01 PB169039BL	4582	17.09	3451	21.28	3078	23.51
02 RW8-SP100-20250724	4635	17.09	3863	21.28	3344	23.51
03 RW8-SP303-20250724	4562	17.09	3546	21.28	3028	23.51
04 PB169039BS	3846	17.09	2873	21.28	2422	23.52
05 PB169039BSD	3782	17.09	2830	21.28	2422	23.52

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.