

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: Alliance
Lab Code: ACE
Lab File ID: BN037547.D
Instrument ID: BNA_N

Contract: TETR06
SDG NO.: Q2696
DFTPP Injection Date: 07/30/2025
DFTPP Injection Time: 08:59

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
68	Less than 2.0% of mass 69	0.0 (0.0) 1
69	Mass 69 relative abundance	38.7
70	Less than 2.0% of mass 69	0.2 (0.6) 1
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.9
365	Greater than 1% of mass 198	4.1
441	Present, but less than mass 443	14.4
442	Greater than 50% of mass 198	90.2
443	15.0 - 24.0% of mass 442	17.3 (19.2) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC0.4	SSTDCCC0.4	BN037548.D	07/30/2025	09:38
PB169039BL	PB169039BL	BN037549.D	07/30/2025	10:14
RW8-SP100-20250724	Q2696-01	BN037550.D	07/30/2025	10:51
RW8-SP303-20250724	Q2696-02	BN037551.D	07/30/2025	11:27
PB169039BS	PB169039BS	BN037555.D	07/30/2025	13:52
PB169039BSD	PB169039BSD	BN037556.D	07/30/2025	14:28
SSTDCCC0.4EC	SSTDCCC0.4	BN037557.D	07/30/2025	15:16