

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q2697

Client ID : SSTDCCC0.4

Date Analyzed: 07/30/2025

Lab File ID: BN037548.D

Time Analyzed: 09:38

Instrument ID: BNA_N

GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	2372	7.717	6287	10.50	3213	14.35
UPPER LIMIT	4744	8.217	12574	10.998	6426	14.845
LOWER LIMIT	1186	7.217	3143.5	9.998	1606.5	13.845
EPA SAMPLE NO.						
01 PB169039BL	2234	7.72	5448	10.50	2592	14.36
02 RW7-SP100-20250724	2145	7.72	5437	10.50	2763	14.36
03 PB169039BS	1872	7.72	4499	10.50	2146	14.35
04 PB169039BSD	1856	7.72	4469	10.50	2108	14.36
05 RW7-SP201-20250724	2519	7.72	6118	10.50	2961	14.36
06 RW7-SP303-20250724	1819	7.72	4480	10.50	2281	14.36

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.