

## Initial and Continuing Calibration Verification

**Client:** Holland Manufacturing Co.

**SDG No.:** Q2698

**Project:** Pre Treatment Plant 2025

**RunNo.:** LB136621

| Analyte                                       | Units | Result | True Value | %<br>Recovery | Acceptance<br>Window (%R) | Analysis<br>Date |
|-----------------------------------------------|-------|--------|------------|---------------|---------------------------|------------------|
| Sample ID: <b>ICV</b><br>Orthophosphate as P  | mg/L  | 0.521  | 0.50       | 104           | 90-110                    | 07/25/2025       |
| Sample ID: <b>CCV1</b><br>Orthophosphate as P | mg/L  | 0.508  | 0.5        | 102           | 90-110                    | 07/25/2025       |
| Sample ID: <b>CCV2</b><br>Orthophosphate as P | mg/L  | 0.496  | 0.5        | 99            | 90-110                    | 07/25/2025       |

## Initial and Continuing Calibration Verification

**Client:** Holland Manufacturing Co.

**SDG No.:** Q2698

**Project:** Pre Treatment Plant 2025

**RunNo.:** LB136623

| Analyte                                     | Units | Result | True Value | %<br>Recovery | Acceptance<br>Window (%R) | Analysis<br>Date |
|---------------------------------------------|-------|--------|------------|---------------|---------------------------|------------------|
| Sample ID: <b>ICV</b><br>Phosphorus, Total  | mg/L  | 0.496  | 0.50       | 99            | 90-110                    | 07/28/2025       |
| Sample ID: <b>CCV1</b><br>Phosphorus, Total | mg/L  | 0.506  | 0.50       | 101           | 90-110                    | 07/28/2025       |
| Sample ID: <b>CCV2</b><br>Phosphorus, Total | mg/L  | 0.490  | 0.50       | 98            | 90-110                    | 07/28/2025       |

## Initial and Continuing Calibration Verification

**Client:** Holland Manufacturing Co.

**SDG No.:** Q2698

**Project:** Pre Treatment Plant 2025

**RunNo.:** LB136664

| Analyte                                | Units | Result | True Value | % Recovery | Acceptance Window (%R) | Analysis Date |
|----------------------------------------|-------|--------|------------|------------|------------------------|---------------|
| Sample ID: <b>ICV1</b><br>Ammonia as N | mg/L  | 1      | 1          | 100        | 90-110                 | 07/31/2025    |
| Sample ID: <b>CCV1</b><br>Ammonia as N | mg/L  | 1      | 1          | 100        | 90-110                 | 07/31/2025    |
| Sample ID: <b>CCV2</b><br>Ammonia as N | mg/L  | 1      | 1          | 100        | 90-110                 | 07/31/2025    |
| Sample ID: <b>CCV3</b><br>Ammonia as N | mg/L  | 0.98   | 1          | 98         | 90-110                 | 07/31/2025    |