

Initial and Continuing Calibration Blank Summary

Client: Holland Manufacturing Co.

SDG No.: Q2698

Project: Pre Treatment Plant 2025

RunNo.: LB136621

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB Orthophosphate as P	mg/L	0.006	0.0250	J	0.0038	0.05	07/25/2025
Sample ID: CCB1 Orthophosphate as P	mg/L	0.004	0.0250	J	0.0038	0.05	07/25/2025
Sample ID: CCB2 Orthophosphate as P	mg/L	0.006	0.0250	J	0.0038	0.05	07/25/2025

Initial and Continuing Calibration Blank Summary

Client: Holland Manufacturing Co.

SDG No.: Q2698

Project: Pre Treatment Plant 2025

RunNo.: LB136623

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB Phosphorus, Total	mg/L	0.006	0.0250	J	0.0045	0.05	07/28/2025
Sample ID: CCB1 Phosphorus, Total	mg/L	0.008	0.0250	J	0.0045	0.05	07/28/2025
Sample ID: CCB2 Phosphorus, Total	mg/L	0.006	0.0250	J	0.0045	0.05	07/28/2025

Initial and Continuing Calibration Blank Summary

Client: Holland Manufacturing Co.

SDG No.: Q2698

Project: Pre Treatment Plant 2025

RunNo.: LB136664

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB1 Ammonia as N	mg/L	< 0.0500	0.0500	U	0.030	0.1	07/31/2025
Sample ID: CCB1 Ammonia as N	mg/L	< 0.0500	0.0500	U	0.030	0.1	07/31/2025
Sample ID: CCB2 Ammonia as N	mg/L	< 0.0500	0.0500	U	0.030	0.1	07/31/2025
Sample ID: CCB3 Ammonia as N	mg/L	< 0.0500	0.0500	U	0.030	0.1	07/31/2025