

Duplicate Sample Summary

Client: Holland Manufacturing Co.	SDG No.: Q2698
Project: Pre Treatment Plant 2025	Sample ID: Q2682-02
Client ID: TOWER-2DUP	Percent Solids for Spike Sample: 0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
TSS	mg/L	+/-5	2.30	J	2.30	J	1	0		07/25/2025

Duplicate Sample Summary

Client: Holland Manufacturing Co.	SDG No.: Q2698
Project: Pre Treatment Plant 2025	Sample ID: Q2698-01
Client ID: EFFLUENTDUP	Percent Solids for Spike Sample: 0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Orthophosphate as P	mg/L	+/-20	0.059		0.059		1	0		07/25/2025
Phosphorus, Total	mg/L	+/-20	0.14		0.14		1	1.47		07/28/2025
BOD5	mg/L	+/-20	5460		5670		1	3.77		07/25/2025

Duplicate Sample Summary

Client: Holland Manufacturing Co.	SDG No.: Q2698
Project: Pre Treatment Plant 2025	Sample ID: Q2698-01
Client ID: EFFLUENTMSD	Percent Solids for Spike Sample: 0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Orthophosphate as P	mg/L	+/-20	0.55		0.54		1	0.37		07/25/2025
Phosphorus, Total	mg/L	+/-20	0.60		0.59		1	1.01		07/28/2025

Duplicate Sample Summary

Client:	Holland Manufacturing Co.	SDG No.:	Q2698
Project:	Pre Treatment Plant 2025	Sample ID:	Q2698-01
Client ID:	EFFLUENTMSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Oil and Grease	mg/L	+/-18	28.2		28.4		1	0.71		07/31/2025

Duplicate Sample Summary

Client: Holland Manufacturing Co.	SDG No.: Q2698
Project: Pre Treatment Plant 2025	Sample ID: Q2702-01
Client ID: MH-7-25-25MSD	Percent Solids for Spike Sample: 0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Oil and Grease	mg/L	+/-18	167		167		1	0.06		07/31/2025

Duplicate Sample Summary

Client: Holland Manufacturing Co.	SDG No.: Q2698
Project: Pre Treatment Plant 2025	Sample ID: Q2730-02
Client ID: 002 35th Ave(july)DUP	Percent Solids for Spike Sample: 0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Ammonia as N	mg/L	+/-20	6.40	OR	6.50	OR	1	2		07/31/2025
Ammonia as N	mg/L	+/-20	5.90	D	6.30	D	5	7		07/31/2025

Duplicate Sample Summary

Client: Holland Manufacturing Co.	SDG No.: Q2698
Project: Pre Treatment Plant 2025	Sample ID: Q2730-02
Client ID: 002 35th Ave(july)MSD	Percent Solids for Spike Sample: 0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Ammonia as N	mg/L	+/-20	7.60	OR	7.60	OR	1	0		07/31/2025