

Laboratory Control Sample Summary

Client: Holland Manufacturing Co.

SDG No.: Q2698

Project: Pre Treatment Plant 2025

Run No.: LB136621

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136621BS							
Orthophosphate as P	mg/L	0.5	0.51		102	1	90-110	07/25/2025

Laboratory Control Sample Summary

Client: Holland Manufacturing Co.

SDG No.: Q2698

Project: Pre Treatment Plant 2025

Run No.: LB136624

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136624BS							
BOD5	mg/L	198	215		109	1	84.6-115.4	07/25/2025

Laboratory Control Sample Summary

Client: Holland Manufacturing Co.

SDG No.: Q2698

Project: Pre Treatment Plant 2025

Run No.: LB136626

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136626BS							
TSS	mg/L	550	532		97	1	90-110	07/25/2025

Laboratory Control Sample Summary

Client: Holland Manufacturing Co.

SDG No.: Q2698

Project: Pre Treatment Plant 2025

Run No.: LB136659

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136659BS							
Oil and Grease	mg/L	20.0	16.9		84	1	78-114	07/31/2025

Laboratory Control Sample Summary

Client: Holland Manufacturing Co.

SDG No.: Q2698

Project: Pre Treatment Plant 2025

Run No.: LB136623

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB169013BS							
Phosphorus, Total	mg/L	0.50	0.49		99	1	90-110	07/28/2025

Laboratory Control Sample Summary

Client: Holland Manufacturing Co.

SDG No.: Q2698

Project: Pre Treatment Plant 2025

Run No.: LB136664

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB169056BS							
Ammonia as N	mg/L	1	1.00		100	1	90-110	07/31/2025