

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tully Environmental, Inc

SDG No.: Q2699

Contract: TULL01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Copper	1020	1000	102	95 - 105	P	07/28/2025	11:38	LB136640
	Iron	3980	4000	100	95 - 105	P	07/28/2025	11:38	LB136640
	Lead	3900	4000	98	95 - 105	P	07/28/2025	11:38	LB136640

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tully Environmental, Inc

SDG No.: Q2699

Contract: TULL01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Copper	22.2	20.0	111	80 - 120	P	07/28/2025	11:53	LB136640
	Iron	90.4	100	90	80 - 120	P	07/28/2025	11:53	LB136640
	Lead	11.8	12.0	98	80 - 120	P	07/28/2025	11:53	LB136640

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tully Environmental, Inc

SDG No.: Q2699

Contract: TULL01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Copper	1250	1250	100	90 - 110	P	07/28/2025	12:29	LB136640
	Iron	4950	5000	99	90 - 110	P	07/28/2025	12:29	LB136640
	Lead	4900	5000	98	90 - 110	P	07/28/2025	12:29	LB136640
CCV02	Copper	1250	1250	100	90 - 110	P	07/28/2025	13:27	LB136640
	Iron	4780	5000	96	90 - 110	P	07/28/2025	13:27	LB136640
	Lead	4910	5000	98	90 - 110	P	07/28/2025	13:27	LB136640
CCV03	Copper	1250	1250	100	90 - 110	P	07/28/2025	14:35	LB136640
	Iron	5130	5000	103	90 - 110	P	07/28/2025	14:35	LB136640
	Lead	4850	5000	97	90 - 110	P	07/28/2025	14:35	LB136640
CCV04	Copper	1270	1250	102	90 - 110	P	07/28/2025	15:40	LB136640
	Iron	5250	5000	105	90 - 110	P	07/28/2025	15:40	LB136640
	Lead	4910	5000	98	90 - 110	P	07/28/2025	15:40	LB136640
CCV05	Copper	1250	1250	100	90 - 110	P	07/28/2025	16:58	LB136640
	Iron	4900	5000	98	90 - 110	P	07/28/2025	16:58	LB136640
	Lead	4790	5000	96	90 - 110	P	07/28/2025	16:58	LB136640