

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: Tully Environmental, Inc

SDG No.: Q2699

Contract: TULL01

Lab Code: ACE

ICS Source: EPA

Instrument ID: P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Copper	20.3	2.0	1015	-18	22	07/28/2025	12:11	LB136640
	Iron	97800	101000	97	85600	116500	07/28/2025	12:11	LB136640
	Lead	-1.40			-12	12	07/28/2025	12:11	LB136640
ICSAB01	Copper	486	511	95	434	588	07/28/2025	12:16	LB136640
	Iron	101000	99300	102	84400	114500	07/28/2025	12:16	LB136640
	Lead	39.0	49.0	80	37	61	07/28/2025	12:16	LB136640
ICSA	Copper	23.2	2.0	1160	-18	22	07/28/2025	12:20	LB136640
	Iron	95500	101000	95	85600	116500	07/28/2025	12:20	LB136640
	Lead	-13.8			-12	12	07/28/2025	12:20	LB136640
ICSAB	Copper	524	511	102	434	588	07/28/2025	12:24	LB136640
	Iron	98700	99300	99	84400	114500	07/28/2025	12:24	LB136640
	Lead	33.5	49.0	68	37	61	07/28/2025	12:24	LB136640