

metals
- 5a -
MATRIX SPIKE SUMMARY

client: <u>Tully Environmental, Inc</u>	level: <u>low</u>	sdg no.: <u>Q2699</u>
contract: <u>TULL01</u>		lab code: <u>ACE</u>
matrix: <u>Water</u>	sample id: <u>Q2699-02</u>	client id: <u>002-35TH-AVE(JUNE)MS</u>
Percent Solids for Sample: <u>NA</u>	Spiked ID: <u>Q2699-02MS</u>	Percent Solids for Spike Sample: <u>NA</u>

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Copper	ug/L	75 - 125	174		33.5		150	93		P
Iron	ug/L	75 - 125	19700		20000		1500	-19		P
Lead	ug/L	75 - 125	465		24.1		500	88		P

metals
- 5a -
MATRIX SPIKE DUPLICATE SUMMARY

client: <u>Tully Environmental, Inc</u>	level: <u>low</u>	sdg no.: <u>Q2699</u>
contract: <u>TULL01</u>		lab code: <u>ACE</u>
matrix: <u>Water</u>	sample id: <u>Q2699-02</u>	client id: <u>002-35TH-AVE(JUNE)MSD</u>
Percent Solids for Sample: <u>NA</u>	Spiked ID: <u>Q2699-02MSD</u>	Percent Solids for Spike Sample: <u>NA</u>

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Copper	ug/L	75 - 125	164		33.5		150	87		P
Iron	ug/L	75 - 125	19200		20000		1500	-50		P
Lead	ug/L	75 - 125	439		24.1		500	83		P