

Instrument ID: **MSVOA_N**

Daily Analysis Runlog For Sequence/QC Batch ID # VN072825

Review By	John Carlone	Review On	7/29/2025 8:39:05 AM
Supervise By	Mahesh Dadoda	Supervise On	7/29/2025 12:10:18 PM
SubDirectory	VN072825	HP Acquire Method	HP Processing Method 624N072825W.M
STD. NAME	STD REF.#		
Tune/Reschk	VP134925		
Initial Calibration Stds	VP134927,VP134928,VP134929,VP134930,VP134931		
CCC	VP134926		
Internal Standard/PEM			
ICV/I.BLK	VP134932		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	BFB	VN087418.D	28 Jul 2025 08:31	JC\MD	Ok
2	VSTDIC005	VN087419.D	28 Jul 2025 09:06	JC\MD	Ok,M
3	VSTDICCC020	VN087420.D	28 Jul 2025 10:06	JC\MD	Ok,M
4	VSTDIC050	VN087421.D	28 Jul 2025 10:27	JC\MD	Ok,M
5	VSTDIC100	VN087422.D	28 Jul 2025 10:48	JC\MD	Ok,M
6	VSTDIC150	VN087423.D	28 Jul 2025 11:09	JC\MD	Ok,M
7	IBLK	VN087424.D	28 Jul 2025 11:31	JC\MD	Ok
8	VSTDICV020	VN087425.D	28 Jul 2025 11:54	JC\MD	Ok,M
9	VN0728WBS01	VN087426.D	28 Jul 2025 12:27	JC\MD	Ok,M
10	VN0728WBSD01	VN087427.D	28 Jul 2025 13:01	JC\MD	Ok,M
11	VN0728WBL01	VN087428.D	28 Jul 2025 13:22	JC\MD	Ok
12	Q2699-01	VN087429.D	28 Jul 2025 13:56	JC\MD	Dilution
13	Q2699-02	VN087430.D	28 Jul 2025 14:17	JC\MD	Ok,M
14	Q2699-01DL	VN087431.D	28 Jul 2025 14:53	JC\MD	Ok
15	IBLK	VN087432.D	28 Jul 2025 15:14	JC\MD	Ok
16	Q2683-07	VN087433.D	28 Jul 2025 15:36	JC\MD	Ok,M
17	Q2683-03	VN087434.D	28 Jul 2025 15:58	JC\MD	Ok,M
18	VSTDCCC020	VN087435.D	28 Jul 2025 16:34	JC\MD	Ok,M

M : Manual Integration