

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2699 Analytical Method: SW624.1  
Client: Tully Environmental, Inc Datafile : VN087426.D

| Lab Sample ID | Parameter     | Spike | Result | Unit | Rec | RPD | Qual | Low | Limits | RPD |
|---------------|---------------|-------|--------|------|-----|-----|------|-----|--------|-----|
|               |               |       |        |      |     |     |      |     | High   |     |
| VN0728WBS01   | Benzene       | 20    | 17.9   | ug/L | 90  |     |      | 65  | 135    |     |
|               | Toluene       | 20    | 18.0   | ug/L | 90  |     |      | 70  | 130    |     |
|               | Ethyl Benzene | 20    | 18.3   | ug/L | 92  |     |      | 60  | 140    |     |
|               | m/p-Xylenes   | 40    | 36.6   | ug/L | 92  |     |      | 87  | 111    |     |
|               | o-Xylene      | 20    | 18.5   | ug/L | 93  |     |      | 87  | 111    |     |