

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: TULL01
 Lab Code: ACE SDG NO.: Q2699
 Lab File ID: VN087420.D Date Analyzed: 07/28/2025
 Instrument ID: MSVOA_N Time Analyzed: 10:06
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	104668	7.80	586753	9.08	490436	11.85
UPPER LIMIT	209336	8.3	1173510	9.582	980872	12.347
LOWER LIMIT	52334	7.3	293377	8.582	245218	11.347
EPA SAMPLE NO.						
001-WILLETS-PT-BLVD (JUNE)	70178	7.80	392636	9.08	348504	11.85
001-WILLETS-PT-BLVD (JUNE) DL	66855	7.79	369763	9.08	330190	11.85
002-35TH-AVE (JUNE)	67508	7.80	376224	9.08	333013	11.85
VN0728WBL01	71572	7.80	407686	9.08	348117	11.85
VN0728WBS01	87324	7.79	492311	9.08	430169	11.85
VN0728WBSD01	82214	7.80	454464	9.08	401032	11.85

IS1 = Bromochloromethane
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.