

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: Alliance

Contract: TULL01

Lab Code: ACE

SDG NO.: Q2699

Lab File ID: VN087418.D

BFB Injection Date: 07/28/2025

Instrument ID: MSVOA_N

BFB Injection Time: 08:31

GC Column: RXI-624 ID: 0.25 (mm)

Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	20.4
75	30.0 - 60.0% of mass 95	44.3
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.4
173	Less than 2.0% of mass 174	0.6 (0.9) 1
174	50.0 - 100.0% of mass 95	72.8
175	5.0 - 9.0% of mass 174	5.3 (7.3) 1
176	95.0 - 101.0% of mass 174	70.8 (97.2) 1
177	5.0 - 9.0% of mass 176	4.9 (7) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

CLIENT ID	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDIC005	VSTDIC005	VN087419.D	07/28/2025	09:06
VSTDIC020	VSTDIC020	VN087420.D	07/28/2025	10:06
VSTDIC050	VSTDIC050	VN087421.D	07/28/2025	10:27
VSTDIC100	VSTDIC100	VN087422.D	07/28/2025	10:48
VSTDIC150	VSTDIC150	VN087423.D	07/28/2025	11:09
VN0728WBS01	VN0728WBS01	VN087426.D	07/28/2025	12:27
VN0728WBSD01	VN0728WBSD01	VN087427.D	07/28/2025	13:01
VN0728WBL01	VN0728WBL01	VN087428.D	07/28/2025	13:22
001-WILLETS-PT-BLVD (JUNE)	Q2699-01	VN087429.D	07/28/2025	13:56
002-35TH-AVE (JUNE)	Q2699-02	VN087430.D	07/28/2025	14:17
001-WILLETS-PT-BLVD (JUNE) DL	Q2699-01DL	VN087431.D	07/28/2025	14:53