

### Laboratory Control Sample Summary

|                 |                                             |                 |          |
|-----------------|---------------------------------------------|-----------------|----------|
| <b>Client:</b>  | Europastry                                  | <b>SDG No.:</b> | Q2702    |
| <b>Project:</b> | MCUA Permit No 14241 - 571 Jersey Ave NB NJ | <b>Run No.:</b> | LB136659 |

| Analyte        | Units      | True Value | Result | Conc. Qualifier | % Recovery | Dilution Factor | Acceptance Limit %R | Analysis Date |
|----------------|------------|------------|--------|-----------------|------------|-----------------|---------------------|---------------|
| Sample ID      | LB136659BS |            |        |                 |            |                 |                     |               |
| Oil and Grease | mg/L       | 20.0       | 16.9   |                 | 84         | 1               | 78-114              | 07/31/2025    |

### Laboratory Control Sample Summary

|                 |                                             |                 |          |
|-----------------|---------------------------------------------|-----------------|----------|
| <b>Client:</b>  | Europastry                                  | <b>SDG No.:</b> | Q2702    |
| <b>Project:</b> | MCUA Permit No 14241 - 571 Jersey Ave NB NJ | <b>Run No.:</b> | LB136660 |

| Analyte   | Units      | True Value | Result | Conc. Qualifier | % Recovery | Dilution Factor | Acceptance Limit %R | Analysis Date |
|-----------|------------|------------|--------|-----------------|------------|-----------------|---------------------|---------------|
| Sample ID | LB136660BS |            |        |                 |            |                 |                     |               |
| TPH       | mg/L       | 20.0       | 17.0   |                 | 85         | 1               | 78-114              | 07/31/2025    |

### Laboratory Control Sample Summary

|                 |                                             |                 |          |
|-----------------|---------------------------------------------|-----------------|----------|
| <b>Client:</b>  | Europastry                                  | <b>SDG No.:</b> | Q2702    |
| <b>Project:</b> | MCUA Permit No 14241 - 571 Jersey Ave NB NJ | <b>Run No.:</b> | LB136660 |

| Analyte   | Units       | True Value | Result | Conc. Qualifier | % Recovery | Dilution Factor | Acceptance Limit %R | Analysis Date |
|-----------|-------------|------------|--------|-----------------|------------|-----------------|---------------------|---------------|
| Sample ID | LB136660BSD |            |        |                 |            |                 |                     |               |
| TPH       | mg/L        | 20.0       | 17.2   |                 | 86         | 1               | 78-114              | 07/31/2025    |