

Matrix Spike Summary

Client:	G Environmental	SDG No.:	Q2704
Project:	Banker	Sample ID:	Q2695-01
Client ID:	RW5-SP100-20250724MS	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Bromide	mg/L	80-120	10.1		0.37	U	10	1	101		07/25/2025
Chloride	mg/L	80-120	13.4	OR	10.8	OR	3	1	87		07/25/2025
Fluoride	mg/L	80-120	2.00		0.11	U	2	1	100		07/25/2025
Nitrite	mg/L	80-120	3.00		0.074	U	3	1	100		07/25/2025
Nitrate	mg/L	80-120	6.30	OR	3.90		2.5	1	96		07/25/2025
Sulfate	mg/L	80-120	16.9		2.60	J	15	1	95		07/25/2025
Orthophosphate as P	mg/L	80-120	5.10		0.34	U	5	1	102		07/25/2025

Matrix Spike Summary

Client:	G Environmental	SDG No.:	Q2704
Project:	Banker	Sample ID:	Q2695-01
Client ID:	RW5-SP100-20250724MSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Bromide	mg/L	80-120	10.0		0.37	U	10	1	100		07/25/2025
Chloride	mg/L	80-120	13.4	OR	10.8	OR	3	1	87		07/25/2025
Fluoride	mg/L	80-120	2.00		0.11	U	2	1	100		07/25/2025
Nitrite	mg/L	80-120	2.90		0.074	U	3	1	97		07/25/2025
Nitrate	mg/L	80-120	6.30	OR	3.90		2.5	1	96		07/25/2025
Sulfate	mg/L	80-120	16.7		2.60	J	15	1	94		07/25/2025
Orthophosphate as P	mg/L	80-120	5.00		0.34	U	5	1	100		07/25/2025