

Laboratory Control Sample Summary

Client: G Environmental

SDG No.: Q2704

Project: Banker

Run No.: LB136635

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136635BSW							
Bromide	mg/L	10	10.2		102	1	90-110	07/25/2025
Chloride	mg/L	3	3.10		103	1	90-110	07/25/2025
Fluoride	mg/L	2	2.00		100	1	90-110	07/25/2025
Nitrite	mg/L	3	3.00		100	1	90-110	07/25/2025
Nitrate	mg/L	2.5	2.50		100	1	90-110	07/25/2025
Sulfate	mg/L	15	15.0		100	1	90-110	07/25/2025
Orthophosphate as P	mg/L	5	5.10		102	1	90-110	07/25/2025