

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: G Environmental

SDG No.: Q2704

Contract: GENV01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Sodium	21000	20000	105	90 - 110	P	07/29/2025	12:20	LB136648

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: G Environmental

SDG No.: Q2704

Contract: GENV01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Sodium	1700	2000	85	80 - 120	P	07/29/2025	13:01	LB136648

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: G Environmental

SDG No.: Q2704

Contract: GENV01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Sodium	26100	25000	104	90 - 110	P	07/29/2025	14:21	LB136648
CCV02	Sodium	22600	25000	91	90 - 110	P	07/29/2025	15:11	LB136648
CCV03	Sodium	23000	25000	92	90 - 110	P	07/29/2025	16:14	LB136648
CCV04	Sodium	24200	25000	97	90 - 110	P	07/29/2025	17:20	LB136648
CCV05	Sodium	23600	25000	94	90 - 110	P	07/29/2025	18:11	LB136648
CCV06	Sodium	21000	25000	84	90 - 110	P	07/29/2025	18:31	LB136648

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: G Environmental

SDG No.: Q2704

Contract: GENV01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Sodium	19100	20000	95	90 - 110	P	07/30/2025	14:11	LB136658

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: G Environmental

SDG No.: Q2704

Contract: GENV01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Sodium	1850	2000	93	80 - 120	P	07/30/2025	14:16	LB136658

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: G Environmental

SDG No.: Q2704

Contract: GENV01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Sodium	23600	25000	95	90 - 110	P	07/30/2025	14:45	LB136658
CCV02	Sodium	22600	25000	90	90 - 110	P	07/30/2025	15:35	LB136658
CCV03	Sodium	23900	25000	96	90 - 110	P	07/30/2025	15:51	LB136658
