



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: G Environmental

SDG No.: Q2704

Contract: GENV01

Lab Code: ACE

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Sodium	1570	2000	78	65 - 135	P	07/29/2025	13:09	LB136648
CRI01	Sodium	1810	2000	90	65 - 135	P	07/30/2025	14:24	LB136658