

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB136648

Review By	jaswal	Review On	7/30/2025 1:09:21 PM
Supervise By	MOHAN	Supervise On	8/1/2025 5:25:28 PM
STD. NAME	STD REF.#		
ICAL Standard	MP86452,MP86453,MP86454,MP86455,MP86456,MP86458		
ICV Standard	MP86459,MP86458		
CCV Standard	MP86462		
ICSA Standard	MP86460,MP86461		
CRI Standard	MP86458		
LCS Standard			
Chk Standard	MP86463,MP86464		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	07/29/25 11:20		Jaswal	OK
2	S1	S1	CAL2	07/29/25 11:24		Jaswal	OK
3	S2	S2	CAL3	07/29/25 11:28		Jaswal	OK
4	S3	S3	CAL4	07/29/25 11:32		Jaswal	OK
5	S4	S4	CAL5	07/29/25 11:37		Jaswal	OK
6	S5	S5	CAL6	07/29/25 11:41		Jaswal	OK
7	ICV01	ICV01	ICV	07/29/25 12:20		Jaswal	OK
8	LLICV01	LLICV01	LLICV	07/29/25 13:01		Jaswal	OK
9	ICB01	ICB01	ICB	07/29/25 13:05		Jaswal	OK
10	CRI01	CRI01	CRDL	07/29/25 13:09		Jaswal	OK
11	ICSA01	ICSA01	ICSA	07/29/25 13:14		Jaswal	OK
12	ICSAB01	ICSAB01	ICSAB	07/29/25 13:20		Jaswal	OK
13	ICSADL	ICSADL	ICSA	07/29/25 13:24		Jaswal	OK
14	ICSABDL	ICSABDL	ICSAB	07/29/25 13:28		Jaswal	OK
15	CCV01	CCV01	CCV	07/29/25 14:21		Jaswal	OK
16	CCB01	CCB01	CCB	07/29/25 14:25		Jaswal	OK
17	PB169029BL	PB169029BL	MB	07/29/25 14:29		Jaswal	OK
18	PB169029BS	PB169029BS	LCS	07/29/25 14:34		Jaswal	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB136648

Review By	jaswal	Review On	7/30/2025 1:09:21 PM
Supervise By	MOHAN	Supervise On	8/1/2025 5:25:28 PM
STD. NAME	STD REF.#		
ICAL Standard	MP86452,MP86453,MP86454,MP86455,MP86456,MP86458		
ICV Standard	MP86459,MP86458		
CCV Standard	MP86462		
ICSA Standard	MP86460,MP86461		
CRI Standard	MP86458		
LCS Standard			
Chk Standard	MP86463,MP86464		

19	Q2700-01	EO-03-072525	SAM	07/29/25 14:38		Jaswal	OK
20	Q2703-01	TP-4	SAM	07/29/25 14:42		Jaswal	OK
21	Q2705-03	FG2A	SAM	07/29/25 14:46		Jaswal	OK
22	Q2705-03DUP	FG2ADUP	DUP	07/29/25 14:50		Jaswal	OK
23	Q2705-03L	FG2AL	SD	07/29/25 14:54		Jaswal	OK
24	Q2705-03MS	FG2AMS	MS	07/29/25 14:59		Jaswal	OK
25	Q2705-03MSD	FG2AMSD	MSD	07/29/25 15:03		Jaswal	OK
26	Q2705-03A	FG2AA	PS	07/29/25 15:07	0.1ml m6004,m6013-10ml sample	Jaswal	OK
27	CCV02	CCV02	CCV	07/29/25 15:11		Jaswal	OK
28	CCB02	CCB02	CCB	07/29/25 15:15		Jaswal	OK
29	Q2705-04	FG2B	SAM	07/29/25 15:19		Jaswal	OK
30	Q2705-05	FG2C	SAM	07/29/25 15:23		Jaswal	OK
31	Q2706-01	RT-5417	SAM	07/29/25 15:27		Jaswal	OK
32	Q2706-03	ETGI-361	SAM	07/29/25 15:31		Jaswal	OK
33	PB169032BL	PB169032BL	MB	07/29/25 15:36		Jaswal	OK
34	Q2696-01	RW8-SP100-2025072	SAM	07/29/25 15:44		Jaswal	OK
35	Q2704-01	MW3	SAM	07/29/25 15:48		Jaswal	OK
36	Q2704-02	MW4	SAM	07/29/25 15:52		Jaswal	OK
37	Q2704-03	MW2D	SAM	07/29/25 15:57		Jaswal	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB136648

Review By	jaswal	Review On	7/30/2025 1:09:21 PM
Supervise By	MOHAN	Supervise On	8/1/2025 5:25:28 PM
STD. NAME	STD REF.#		
ICAL Standard	MP86452,MP86453,MP86454,MP86455,MP86456,MP86458		
ICV Standard	MP86459,MP86458		
CCV Standard	MP86462		
ICSA Standard	MP86460,MP86461		
CRI Standard	MP86458		
LCS Standard			
Chk Standard	MP86463,MP86464		

38	PB169032BS	PB169032BS	LCS	07/29/25 16:06		Jaswal	OK
39	CCV03	CCV03	CCV	07/29/25 16:14		Jaswal	OK
40	CCB03	CCB03	CCB	07/29/25 16:20		Jaswal	OK
41	Q2696-02	RW8-SP303-2025072	SAM	07/29/25 16:24	NOT USE	Jaswal	Not Ok
42	Q2696-02DUP	RW8-SP303-2025072	DUP	07/29/25 16:28	NOT USE	Jaswal	Not Ok
43	Q2696-02L	RW8-SP303-2025072	SD	07/29/25 16:32	NOT USE	Jaswal	Not Ok
44	Q2696-02MS	RW8-SP303-2025072	MS	07/29/25 16:36	NOT USE	Jaswal	Not Ok
45	Q2696-02MSD	RW8-SP303-2025072	MSD	07/29/25 16:41	NOT USE	Jaswal	Not Ok
46	Q2696-02A	RW8-SP303-2025072	PS	07/29/25 16:45	NOT USE(0.1ml m6004,m6013-10ml sample	Jaswal	Not Ok
47	PB169048BL	PB169048BL	MB	07/29/25 16:55		Jaswal	OK
48	PB169048BS	PB169048BS	LCS	07/29/25 17:08		Jaswal	OK
49	Q2417-02DL	COMPDL	SAM	07/29/25 17:12	Straight 5x for Dilution for all elements	Jaswal	OK
50	Q2417-02DUPDL	COMPDUPL	DUP	07/29/25 17:16	Straight 5x for Dilution for all elements	Jaswal	OK
51	CCV04	CCV04	CCV	07/29/25 17:20		Jaswal	OK
52	CCB04	CCB04	CCB	07/29/25 17:24		Jaswal	OK
53	Q2417-02LDL	COMPLDL	SD	07/29/25 17:29	Straight 25x for Dilution for all elements	Jaswal	OK
54	Q2417-02MSDL	COMPMSDL	MS	07/29/25 17:33	Straight 5x for Dilution for all elements	Jaswal	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB136648

Review By	jaswal	Review On	7/30/2025 1:09:21 PM
Supervise By	MOHAN	Supervise On	8/1/2025 5:25:28 PM

STD. NAME	STD REF.#
ICAL Standard	MP86452,MP86453,MP86454,MP86455,MP86456,MP86458
ICV Standard	MP86459,MP86458
CCV Standard	MP86462
ICSA Standard	MP86460,MP86461
CRI Standard	MP86458
LCS Standard	
Chk Standard	MP86463,MP86464

55	Q2417-02MSDDL	COMPMSDDL	MSD	07/29/25 17:37	Straight 5x for Dilution for all elements	Jaswal	OK
56	Q2417-02ADL	COMPADL	PS	07/29/25 17:41	Straight 5x for Dilution for all elements(0.1ml m6004,m6013-10ml sample before deiltion)	Jaswal	OK
57	Q2706-02	RT-5417	SAM	07/29/25 17:45		Jaswal	OK
58	Q2706-04	ETGI-361	SAM	07/29/25 17:50		Jaswal	OK
59	Q2710-01	JC-03-07282025	SAM	07/29/25 17:54	NOT USE	Jaswal	Not Ok
60	Q2710-01DUP	JC-03-07282025DUP	DUP	07/29/25 17:58	NOT USE	Jaswal	Not Ok
61	Q2710-01L	JC-03-07282025L	SD	07/29/25 18:02	NOT USE	Jaswal	Not Ok
62	Q2710-01MS	JC-03-07282025MS	MS	07/29/25 18:07	NOT USE	Jaswal	Not Ok
63	CCV05	CCV05	CCV	07/29/25 18:11		Jaswal	OK
64	CCB05	CCB05	CCB	07/29/25 18:15		Jaswal	OK
65	Q2710-01MSD	JC-03-07282025MSD	MSD	07/29/25 18:19	CCV fail for Be,Na	Jaswal	Not Ok
66	Q2710-01A	JC-03-07282025A	PS	07/29/25 18:23	CCV fail for Be,Na	Jaswal	Not Ok
67	CCV06	CCV06	CCV	07/29/25 18:31	CCV fail for Be,Na(0.1ml m6004,m6013-10ml sample)	Jaswal	OK
68	CCB06	CCB06	CCB	07/29/25 18:36		Jaswal	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB136688

Review By	Janvi	Review On	8/5/2025 10:09:04 AM
Supervise By	jaswal	Supervise On	8/6/2025 1:49:43 AM
STD. NAME	STD REF.#		
ICAL Standard	MP86452,MP86453,MP86454,MP86455,MP86456,MP86458		
ICV Standard	MP86459		
CCV Standard	MP86462		
ICSA Standard	MP86460,MP86461		
CRI Standard	MP86458		
LCS Standard			
Chk Standard	MP86463,MP86464		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	08/01/25 11:18		Jaswal	OK
2	S1	S1	CAL2	08/01/25 11:22		Jaswal	OK
3	S2	S2	CAL3	08/01/25 11:26		Jaswal	OK
4	S3	S3	CAL4	08/01/25 11:30		Jaswal	OK
5	S4	S4	CAL5	08/01/25 11:34		Jaswal	OK
6	S5	S5	CAL6	08/01/25 11:38		Jaswal	OK
7	ICV01	ICV01	ICV	08/01/25 11:43		Jaswal	OK
8	LLICV01	LLICV01	LLICV	08/01/25 11:54		Jaswal	OK
9	ICB01	ICB01	ICB	08/01/25 11:59		Jaswal	OK
10	CRI01	CRI01	CRDL	08/01/25 12:04		Jaswal	OK
11	ICSA01	ICSA01	ICSA	08/01/25 12:09		Jaswal	OK
12	ICSAB01	ICSAB01	ICSAB	08/01/25 12:14		Jaswal	OK
13	ICSADL	ICSADL	ICSA	08/01/25 12:18		Jaswal	OK
14	ICSABDL	ICSABDL	ICSAB	08/01/25 12:22		Jaswal	OK
15	CCV01	CCV01	CCV	08/01/25 12:26		Jaswal	OK
16	CCB01	CCB01	CCB	08/01/25 12:30		Jaswal	OK
17	PB169062BL	PB169062BL	MB	08/01/25 12:35		Jaswal	OK
18	PB169062BS	PB169062BS	LCS	08/01/25 12:39		Jaswal	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB136688

Review By	Janvi	Review On	8/5/2025 10:09:04 AM
Supervise By	jaswal	Supervise On	8/6/2025 1:49:43 AM
STD. NAME	STD REF.#		
ICAL Standard	MP86452,MP86453,MP86454,MP86455,MP86456,MP86458		
ICV Standard	MP86459		
CCV Standard	MP86462		
ICSA Standard	MP86460,MP86461		
CRI Standard	MP86458		
LCS Standard			
Chk Standard	MP86463,MP86464		

19	Q2705-02	FG1B	SAM	08/01/25 12:43		Jaswal	OK
20	Q2705-02DUP	FG1BDUP	DUP	08/01/25 12:47		Jaswal	OK
21	Q2705-02L	FG1BL	SD	08/01/25 12:51		Jaswal	OK
22	Q2705-02MS	FG1BMS	MS	08/01/25 12:55		Jaswal	OK
23	Q2705-02MSD	FG1BMSD	MSD	08/01/25 12:59		Jaswal	OK
24	Q2705-02A	FG1BA	PS	08/01/25 13:03	0.1ml m6004,m6013-10ml sample	Jaswal	OK
25	Q2726-01	WBR-PBO	SAM	08/01/25 13:08		Jaswal	OK
26	Q2728-01	7211931-VNJ240	SAM	08/01/25 13:12		Jaswal	OK
27	CCV02	CCV02	CCV	08/01/25 13:16		Jaswal	OK
28	CCB02	CCB02	CCB	08/01/25 13:22		Jaswal	OK
29	PB169063BL	PB169063BL	MB	08/01/25 13:29		Jaswal	OK
30	PB169063BS	PB169063BS	LCS	08/01/25 13:33		Jaswal	OK
31	PB169060TB	PB169060TB	MB	08/01/25 13:37		Jaswal	OK
32	Q2728-02	7211931-VNJ240	SAM	08/01/25 13:42		Jaswal	OK
33	Q2728-02DUP	7211931-VNJ240DUP	DUP	08/01/25 13:46		Jaswal	OK
34	Q2728-02L	7211931-VNJ240L	SD	08/01/25 13:50		Jaswal	OK
35	Q2728-02MS	7211931-VNJ240MS	MS	08/01/25 13:55		Jaswal	OK
36	Q2728-02MSD	7211931-VNJ240MSD	MSD	08/01/25 13:59		Jaswal	OK
37	Q2728-02A	7211931-VNJ240A	PS	08/01/25 14:03	0.1ml m6004,m6013-10ml sample	Jaswal	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB136688

Review By	Janvi	Review On	8/5/2025 10:09:04 AM
Supervise By	jaswal	Supervise On	8/6/2025 1:49:43 AM
STD. NAME	STD REF.#		
ICAL Standard	MP86452,MP86453,MP86454,MP86455,MP86456,MP86458		
ICV Standard	MP86459		
CCV Standard	MP86462		
ICSA Standard	MP86460,MP86461		
CRI Standard	MP86458		
LCS Standard			
Chk Standard	MP86463,MP86464		

38	CCV03	CCV03	CCV	08/01/25 14:07		Jaswal	OK
39	CCB03	CCB03	CCB	08/01/25 14:16		Jaswal	OK
40	Q1168-01	LOD-MDL-SOIL-01-Q	SAM	08/01/25 14:42		Jaswal	OK
41	Q1168-03	MDL-SOIL-03-QT1-20	SAM	08/01/25 14:47		Jaswal	OK
42	Q1168-07	LOD-MDL-WATER-01	SAM	08/01/25 14:51		Jaswal	OK
43	Q1168-09	MDL-WATER-03-QT1	SAM	08/01/25 14:55		Jaswal	OK
44	Q1168-10	LOD-MDL-WATER-04	SAM	08/01/25 14:59		Jaswal	OK
45	Q1168-12	MDL-WATER-06-QT1	SAM	08/01/25 15:04		Jaswal	OK
46	Q2126-01	LOD-MDL-SOIL-03-Q	SAM	08/01/25 15:08		Jaswal	OK
47	Q2126-03	MDL-SOIL-03-QT2-20	SAM	08/01/25 15:12		Jaswal	OK
48	Q2126-07	LOD-MDL-WATER-01	SAM	08/01/25 15:17		Jaswal	OK
49	Q2126-09	MDL-WATER-03-QT2	SAM	08/01/25 15:21		Jaswal	OK
50	CCV04	CCV04	CCV	08/01/25 15:32		Jaswal	OK
51	CCB04	CCB04	CCB	08/01/25 15:36		Jaswal	OK
52	Q2126-12	MDL-WATER-06-QT2	SAM	08/01/25 15:40		Jaswal	OK
53	Q2126-10	LOD-MDL-WATER-04	SAM	08/01/25 15:45		Jaswal	OK
54	PB167019BS	PB167019BS	LCS	08/01/25 15:53		Jaswal	OK
55	PB167019BL	PB167019BL	MB	08/01/25 16:03		Jaswal	OK
56	PB166198BL	PB166198BL	MB	08/01/25 16:07		Jaswal	OK
57	PB166198BS	PB166198BS	LCS	08/01/25 16:15		Jaswal	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB136688

Review By	Janvi	Review On	8/5/2025 10:09:04 AM
Supervise By	jaswal	Supervise On	8/6/2025 1:49:43 AM
STD. NAME	STD REF.#		
ICAL Standard	MP86452,MP86453,MP86454,MP86455,MP86456,MP86458		
ICV Standard	MP86459		
CCV Standard	MP86462		
ICSA Standard	MP86460,MP86461		
CRI Standard	MP86458		
LCS Standard			
Chk Standard	MP86463,MP86464		

58	PB167652BL	PB167652BL	MB	08/01/25 16:19		Jaswal	OK
59	PB167652BS	PB167652BS	LCS	08/01/25 16:23		Jaswal	OK
60	PBW	PBW	MB	08/01/25 16:27		Jaswal	OK
61	LCSW	LCSW	LCS	08/01/25 16:31		Jaswal	OK
62	CCV05	CCV05	CCV	08/01/25 16:35		Jaswal	OK
63	CCB05	CCB05	CCB	08/01/25 16:40		Jaswal	OK
64	PBS	PBS	MB	08/01/25 16:44		Jaswal	OK
65	LCSS	LCSS	LCS	08/01/25 16:48		Jaswal	OK
66	PB168166BL	PB168166BL	MB	08/01/25 16:52		Jaswal	OK
67	PB168166BS	PB168166BS	LCS	08/01/25 16:57		Jaswal	OK
68	CCV06	CCV06	CCV	08/01/25 17:01		Jaswal	OK
69	CCB06	CCB06	CCB	08/01/25 17:05		Jaswal	OK