

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: G Environmental

SDG No.: Q2705

Contract: GENV01

Lab Code: ACE

ICS Source: EPA

Instrument ID: P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Arsenic	7.46			-20	20	07/29/2025	13:14	LB136648
ICSAB01	Arsenic	104	104	100	88.4	120	07/29/2025	13:20	LB136648
ICSA01	Arsenic	8.00			-20	20	08/01/2025	12:09	LB136688
ICSAB01	Arsenic	104	104	100	88.4	120	08/01/2025	12:14	LB136688