

Continuing Calibration Report for SequenceID : FC072825AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC069543.D

Aliphatic C9-C12	8377980.000	60.000	3.303	6.602	139633.000	138999.501	-0.456
Aliphatic C12-C16	5704222.000	40.000	6.603	10.006	142605.550	144073.407	1.019
Aliphatic C16-C21	8072340.000	60.000	10.007	13.376	134539.000	135996.946	1.072
Aliphatic C21-C28	9651896.000	80.000	13.377	17.043	120648.700	118989.055	-1.395
Aliphatic C28-C40	11839341.000	120.000	17.044	22.029	98661.175	89883.162	-9.766
Aliphatic EPH	43645779.000	360.000	3.303	22.029	121238.275	118243.964	-2.532