

Continuing Calibration Report for SequenceID : FC072825AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC069555.D

Aliphatic C9-C12	8223559.000	60.000	3.303	6.602	137059.317	138999.501	1.396
Aliphatic C12-C16	5506419.000	40.000	6.603	10.006	137660.475	144073.407	4.451
Aliphatic C16-C21	7854249.000	60.000	10.007	13.376	130904.150	135996.946	3.745
Aliphatic C21-C28	9541715.000	80.000	13.377	17.043	119271.438	118989.055	-0.237
Aliphatic C28-C40	11960903.000	120.000	17.044	22.029	99674.192	89883.162	-10.893
Aliphatic EPH	43086845.000	360.000	3.303	22.029	119685.681	118243.964	-1.219